



JORDAN ARMED FORCES



The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 1	Wheel Chair, Adult	Qty. (175)
--------	--------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		

[Handwritten signature]



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Adult size wheel-chair (seat width 42-44 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 2	Wheel Chair, Extra Wide Adult	Qty. (50)
--------	-------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Extra wide Adult size wheel-chair (seat width 56-60 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 3	Wheel Chair, Children	Qty. (50)
--------	-----------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Children size wheel-chair (seat width 36-38 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 4	Wheel Chair, Wide Adult	Qty. (175)
--------	-------------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	wide Adult size wheel-chair (seat width 46-48 cm)		



JORDAN ARMED FORCES

The DIRECTION OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 5	Wheel Chair, Narrow Adult	Qty. (40)
--------	---------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Narrow Adult size wheel-chair (seat width 38-40cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 6	WHEELCHAIR, WIDE ADULT (FULLY RECLINING BACK)	Qty. (20)
--------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair wide Adult size (seat width 46-48 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 7	WHEELCHAIR, ADULT(FULLY RECLINING BACK)	Qty. (30)
--------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Adult size (seat width 42-44 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 8	WHEELCHAIR,NARRW ADULT(FULLY RECLINING BACK)	Qty. (10)
--------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Narrow Adult size (seat width 38-40 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 9	WHEELCHAIR, CHILDREN (FULLY RECLINING BACK)	Qty. (40)
--------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair children size (seat width 36-38cm)		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 10	WHEELCHAIR FOR CHILDREN WITH CEREBRAL PALSY(36cm)	Qty. (10)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction aluminium frame.		
2	Fully reclining back (90-180) degree with head support wheel-chair (seat width 36 cm)		
3	Adjustable seat angle.		
4	Detachable desk arms rest		
5	Pneumatic tired wheels for outdoor use (pump & English key are essential)		
6	Frame made for durable, anti-corrosion, easy to clean		
7	Durable front caster solid or pneumatic to be 360 pivot type, with locking mechanism		
8	Adjustable height head rest (for head stability)		
9	Adjustable height foot plates		
10	Padded abductor for thigh separation		
11	Padding material should be durable, provides optimum comfort, easy to clean and disinfect.		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
12	The wheelchair should have foldable seat and backrest.		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 11	WHEELCHAIR FOR CHILDREN WITH CEREBRAL PALSY(30cm)	Qty. (35)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction aluminium frame.		
2	Fully reclining back (90-180) degree with head support wheel-chair (seat width 30 cm)		
3	Adjustable seat angle.		
4	Detachable desk arms rest		
5	Pneumatic tired with wheels for outdoor use (pump & English key are essential)		
6	Frame made for durable, anti-corrosion, easy to clean		
7	Durable front caster solid or pneumatic to be 360 pivot type, with locking mechanism		
8	Adjustable height head rest (for head stability)		
9	Adjustable height foot plates		
10	Padded abductor for thigh separation		
11	Padding material should be durable, provides optimum comfort, easy to clean and disinfect.		



JORDAN ARMED FORCES

**The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY**



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
12	The wheelchair should have foldable seat and backrest.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 12	Commode chair wide adult size (seat 46-48 cm)	Qty. (50)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Wide Adult size self-propelled shower & commode chair (seat width 46-48 cm).		
2	Flip up arm rest.		
3	Weight capacity \geq 135 kg		
4	Pneumatic tired nylon wheels with hand rims with(pump & English) or rigid type		
5	Solid front castors wheels with lock		
6	Detachable / swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 13	Commode chair adult size (seat 42-44cm)	Qty. (50)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size self-propelled shower & commode chair (seat width 42-44 cm).		
2	Flip up arm rest.		
3	Weight capacity \geq 100 kg		
4	Pneumatic tired nylon wheels with hand rims with(pump & English) or rigid type		
5	Solid front castors wheels with lock		
6	Detachable / swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES

The DIRECTION OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 14	Walker, wide adult	Qty. (75)
---------	--------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Wide adult size walker. Approximately 95 cm height 55cm width		
2	Adjustable height. 30cm – 40cm		
3	Aluminium bar.		
4	Foam hand grip.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 15	Walker, adult	Qty. (75)
---------	---------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Normal adult size walker Approximately 95 cm height 48cm width		
2	Adjustable height. 30cm – 40cm		
3	Aluminium bar.		
4	Foam hand grip.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 16	K-Walker, large	Qty. (25)
---------	-----------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 70cm – 80cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 17	K-Walker, small	Qty. (35)
---------	-----------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 2 to 7.		
3	Adjustable height. 55cm – 65cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES



The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 18	Two fronts Walker style children size (2-7)	Qty. (20)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 2 to 7.		
3	Adjustable height. 45cm – 55cm		
4	Aluminium bars.		
5	Padded hand grip.		
6	Two front wheels		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 19	Two fronts Walker style children size (7-12)	Qty. (10)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 80cm – 90cm		
4	Aluminium bars.		
5	padded hand grip.		
6	Two front wheels		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 20	Standing Frame style (supine/prone) medium size	Qty. (30)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size standing frame.		
2	For children 7-15 years.		
3	Manual height adjustment of top.		
4	lift up to around 140kgs patient from sitting to standing position.		
5	Dimensions of top please specify with raised rim.		
6	Cut-out size is approximately		
7	Padded knee.		
8	Chest and back supports holds patient in standing position.		
9	Stainless steel frame or equivalent		
10	Retractable locking swivel casters.		
11	Frame base is please specify.		
12	Padded Head support .		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 21	Wheelchair pure gel medical cushion (43*41*3 cm)	Qty. (50)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm)		
3	Cushion for adult size wheel-chair (43cm width * 41 length)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 22	Wheelchair pure gel medical cushion (40*40*3 cm)	Qty. (50)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm.)		
2	Cushion for narrow adult size wheel-chair (40cm width * 40 cm length).		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 23	Wheelchair pure gel medical cushion (36*28*3 cm)	Qty. (10)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm.)		
3	Cushion for children size wheel-chair (36cm width * 28cm length).		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 24	Wheelchair pure gel medical cushion (56*60*3 cm)	Qty. (10)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm.)		
3	Cushion for extra wide adult size wheel-chair (56cm width-60cm length).		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 25	Elbow crutches (Tall adult)	Qty. (10)
---------	-------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Tall Adult size stick.130cm – 150cm height		
2	Strong, lightweight aluminium crutch.		
	Adjustable height.		
4	Padding hand grip		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 26	Elbow crutches (Adult)	Qty. (30)
---------	--------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size stick. 110 – 130 cm height		
2	Strong, lightweight aluminium crutch.		
3	Adjustable height.		
4	Padding hand grip		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 27	Patient stick , Quadruped cane	Qty. (30)
---------	--------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size stick.		
2	Aluminium shaft with chrome plated steel base.		
3	Adjustable quadruped stick.		
4	Moulded handle and plastic hand grip.		
5	Adjustable height approx. (70cm_95cm).		
6	Easily reverses for right or left hand.		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 28	Commode chair children size (seat 36-38 cm)	Qty. (5)
---------	---	------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size self-propelled shower & commode chair (seat width 36-38cm).		
2	Flip up arm rest.		
3	Weight capacity \geq 80 kg		
4	Pneumatic tired nylon wheels with hand rims(pump & English) or rigid type		
5	Solid front castors wheels with lock		
6	Detachable/swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		

	Minimum Requirements	Compliance (Y/N), Notes	Document Page No.
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 29	K-Walker, medium	Qty. (50)
---------	------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 60 -70 cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 30	Standing Frame style (supine/prone) small size	Qty. (60)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size standing frame.		
2	For children 2-7years.		
3	Manual height adjustment of top .		
4	Unit can lift up to around 140kgs patient from sitting to standing position.		
5	Padded head support .		
6	Dimensions of top with raised rim please specify all dimensions.		
7	Cut-out size is approximately		
8	Padded knee		
9	Chest and back supports holds patient in standing position.		
10	Stainless steel frame or equivalent		
11	Retractable locking swivel casters.		
12	Frame base please specify		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

SPECIAL TERMS

- *Offers not complying with any of the special terms or the technical specifications shall be considered non-conforming with tender requirements.*
- *Any vendor providing FORGED documents shall be disqualified from the current tender and banned from participating in any future RMS tenders.*
 1. *Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first third of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period shall be rejected.*
 2. *Technical offers must include clear original technical brochures/catalogues for all offered items.*
 3. *Offers must include fully detailed technical offers and compliance sheets as a soft copy (either Microsoft office or Microsoft excel format) in addition to a hard copy, mentioning the exact model/catalogue number and country of origin of the offered item(s), full technical description/specifications and any accessories or options included in the offer.*
 4. *Compliance sheets must be as per the tabular format of the technical specifications in the tender documents, listing the required specifications on one column and a Yes or NO response to each point in the adjacent column, with reference to page and line numbers in the relevant technical brochure. Offers not complying with this term shall be rejected.*
 5. *Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores and bringing the equipment into service.*
 6. *Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).*



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



7. The supplier must furnish DRMS with a guarantee stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded equipment valid for twelvemonths from the date of final acceptance of the equipment by DRMS .
8. a. DRMS has the right to increase the awarded quantities by a Percentage not exceeding 35% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.
- b. DRMS has the right to decrease the awarded quantity by a Percentage not exceeding 50% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.

Q. For offers submitted in Jordanian dinars , payment will be either by wire transfer or by cheque after final acceptance of goods .Any other way of payment will be rejected .