

DIRECTORATE ROYAL MEDICAL SERVICES

Quality Control

TENDER NO P37-2022-20

Item No	Supplier No	REQ. QTY	UNIT	Description
---------	-------------	----------	------	-------------

09_9 Quality Control Department

09_9 Internal Quality Control

1	09_9_01_001	720	Vial	Lyophilised comprehensive control for monitoring Precision of Laboratory test for Immunoassay analytes for low level
2	09_9_01_0011	720	Vial	Lyophilised Comprehensive Control for monitoring precision for Immunoassay control for all analytes for normal level.
3	09_9_01_0012	720	Vial	Lyophilised Comprehensive Control for monitoring precision for Immunoassay control for all analytes for high level.
4	09_9_01_002	180	Vial	Lyophilised Comprehensive Control for monitoring precision of Laboratory test for Tumor Markers analytes for Low Level
5	09_9_01_0021	180	Vial	Lyophilised Comprehensive control for monitoring precision of Laboratory analytes for Normal Level for Tumor Marker.
6	09_9_01_0022	180	Vial	Lyophilised Comprehensive control for monitoring precision for Tumor Marker level high
7	09_9_01_003	720	Vial	Lyophilised Assayed chemistry control Normal level for monitoring precision of laboratory analytes.
8	09_9_01_004	720	Vial	Lyophilised comprehensive control for monitoring precision for all chemistry analytes for abnormal level .
9	09_9_01_006	240	Vial	Lyophilised Diabetes controls ,human whole blood for monitoring hemoglobin HA1C ,Multi-Level.

09_9 External Quality Assurance Program

10	09_9_01_008	14	KIT	Clinical Chemistry Program /External program .
11	09_9_01_010	2	KIT	Immunoassays Program /External program
12	09_9_01_015	3	KIT	Heamatology program /External program.
13	09_9_01_016	2	KIT	Coagulation Program /External program.
14	09_9_01_018	1	KIT	Hemoglobin Program/ External program .

DIRECTORATE ROYAL MEDICAL SERVICES

Quality Control

TENDER NO P37-2022-20

Item No	Stock No	Req Qty	Unit	Description
09_9 Quality Control Department				
09_9 Internal Quality Control				
15	09_9_01_026	480	Vial	Lypholised control for monitoring precision for coagulatoin analytes for Normal level . (Normal level).
16	09_9_01_0261	480	Vial	Lypholised control for monitoring precision for coagulation analytes for abnormal level
17	09_9_01_032	4	KIT	Liquid homocystein human source material for monitoring precision for multi-levels.
18	09_9_01_033	4	KIT	Liquid for Ethanol/Ammonia control level -1 , prolonged open vial stability
19	09_9_01_0331	4	KIT	Liquid for Ethanol/Ammonia control level -2, prolonged open vial stability
09_9Linearity FD General chemistry				
20	09_9_01_0391	10	KIT	general chemistry(GC1)
21	09_9_01_0392	11	KIT	general chemistry(GC3)
22	09_9_01_0394	8	KIT	Lipoproteins (LP)
23	09_9_01_0395	8	KIT	Linearity FD General chemistry(GC2)
24	09_9_01_0396	3	KIT	intended for in vitro diagnostic use in the quantitative determination of linearity calibration verification nd verification of reputable range in automated ,semi automated and manual instrumented system for the following analytes creatine kinase ,MB(CKMB) and myoglobin (MYO),Eech test consist one bottle each of level 1 throgh5,each bottle contain 3ml the exists alinear relationship amany level 1-5 instrument, beckman coulter au cobas
25	09_9_01_040	5	KIT	Linearity FD HbA1C 5X0.5 ml
09_9Linearity FD immunoassay and tumer marker				
26	09_9_01_041	4	KIT	Linearity FD Immunoassay kit 10X5 ml
27	09_9_01_042	4	KIT	Linearity FD Tumor Markers kit 5X1 ml

الشروط الخاصة التسلسل

- 1 - Item(1,2,3,4,5,6,7,8,9,15,16)to be lyophilized Human Source material ,Item (17,18,19) to be liquid, and all these items to be with long shelf life not less than ONE Year,Assayed values for automated instrument.The company should provide an on -line And software programe to enter the results of the Internal quality control(Third part)that improve the lab analytical perfomance and receive monthly report with peer group statistics and provide analytical goals and westgards rules and also provide the choice of use the total allowable error to set up the quality spicification ...
- 2 - Item (10,11,13.14); An External programe each cycle should have a dated programe and a dated samples to evaluate the test we need for our lab , and should be (Lypholised) The External Programe should be an online programe to enter the results and to monitor the accuracy and the performance of instrument during the year .
- 3 - Item (12); An External programe each cycle should have a dated programe and a dated samples to evaluate the test we need for our lab , and should be (whole blood) The External Programe should be an online programe to enter the results and to monitor the accuracy and the performance of instrument during the year . And a full report should be provided to all the participant .
- 4 - Item (20,21,22,23,24,25,26,27);for linearity and calibration verification products,they should be liquid,ready to use,multi-level sets,the shipping on our schedule,expert tecchnical supports,and global network and best practice goals.

الشروط الخاصة للعطاء رقم Quality Control P37-2022-20

الشروط الخاصة التسلسل

5	- Every Item and its levels should has the same lot Number for Each Consigment.
6	- All the tender items must be delivered to princess Eman Research &Laboratory Scietific Center
7	- Dispatched / delivered goods must show description of goods , expiry date , lot No ., Storage conditions , stamped " Sold to RMS " tender (P37-2022-20),and any other necessary information .
8	- Goods should be delivered under the same storage conditions that comply with their nature .,storage conditions must be mentioned clearly on all relative documents.
9	- Any products derived or containing materials of human origin must be clearly indicated and supplied with a certificate &/ or official documents stating that the products are devoid of blood borne infectious agents : i. e HbsAg, HCV,HIV, etc...
10	- All items should be evaluated previously at RMS/QC laboratory according to DRMS REGULATION
11	- All goods should retain at least the MSL mentioned per Each item Any item not Complying with this term upon receipt should be accompanied with a conformation that you accept to replace any remaining unused quantities at your expenses after the expiry date ,said conformation is subject to the approval of the director of the Royal Medical Services and will Incur afine ,which will be decided later by DRMS according to the loss that this discrepancy with terms of the tender has caused.

بسم الله الرحمن الرحيم

القيادة العامة للقوات المسلحة الاردنية

مديرية الخدمات الطبية الملكية

التزويد

صفحة رقم 3 - 3

الشروط الخاصة للعتاء رقم Quality Control P37-2022-20

الشروط الخاصة التسلسل

- 12 - For offers submitted in Jordanian dinars payment will be either by wire transfer or by cheque to be paid after receiving goods ,Any other way of payment will not be accepted and the offers will be rejected by the purchase committee.