

بسم الله الرحمن الرحيم

محضر اجتماع العطاء رقم (ش/٢٠٢٤/٣٧/ACD)

-اجتمعت لجنة الدراسة بتاريخ ٢٠٢٤/٧/٣ لتوضيح الاستفسارات الواردة بمذكرتي شعبية
المشتريات المركزية ارقام (ش/٢٠٢٤/٣٧/ACD/٣٠٠٨ تاريخ ٢٠٢٤/٦/٢٥)
(ش/٢٠٢٤/٣٧/ACD/٣٠٩١ تاريخ ٢٠٢٤/٧/١) وقررت مايلي:-

-فيما يخص تحديد النسب والتوزيع على الشركات يتم تحديدها من قبل لجنة الشراء عند فتح
العروض المقدمة من الشركات.

-يعتبر (HIGH RANGE) وحدة واحدة من حيث المواد وعدد الاجهزة وكذلك (MID
RANGE) وحدة واحدة من حيث المواد وعدد الاجهزة.

-فيما يخص شرط التجريب تعتبر جميع الاجهزة الموجودة في الخدمات الطبية الملكية حاليا
وسابقا مجربة ومعتمدة لدى الخدمات الطبية الملكية لجميع الفحوصات والمعاليل المصنعة من
نفس الشركة وبغض النظر عن سعة الكت وطريقة التغليف لجميع موديلات الاجهزة التي من
نفس الشركة الصانعة وجميع تحديثاتها وبخلاف ذلك فإن جميع الاجهزة الغير متوافدة في
الخدمات الطبية الملكية ولم يتم شراء معاليل لها سابقا تعتبر غير مجربة ولاصانع من تجهيلها
مستقبلا في حال توافر فيها شروط التجريب.

-فيما يخص أنظمة التشغيل المفتوحة OPEN SYSTEM و الفحوصات اليدوية يشترط التجريب
في المواد كونه لايشترط تشغيلها على اجهزة الشركة الصانعة.

-تحتفظ الخدمات الطبية الملكية بحق الغاء ورفض وتفرير اي/المواد والكتات المجربة في
الخدمات الطبية في حال حدوث اي خلل او ورود شكوى.



الرقم : ش/٣٧/٢٠٢٤/ACD
التاريخ : ١٤٤٥/ /
: ٢٠٢٤/١٦/٢٥

السادة: المناقصون
الموضوع : عطاء ش/٣٧/٢٠٢٤/ACD

تحية وبعد،،،

- الاشاره : العطاء رقم ش/٣٧/٢٠٢٤/ACD

١. يعدل الشرط رقم ٧-ح ليصبح كالتالي :

" لا مانع من تقديم اجهزة من موديلات احدث للاجهزة المناقص عليها ومن نفس فئة الاجهزة المحتسبة الموجودة في الخدمات الطبية الملكية علماً انه لا مانع بأن تكون المواد التشغيلية هي نفسها للاجهزة القديمة او المواد تشغيلية للاجهزة الاحدث "

٢. للعلم لطفاً .

واقبلوا فائق الاحترام،،،،،

ع/العميد الطبيب
مدير عام الخدمات الطبية الملكية
علاء الدين خالد أبو رمان

محضر اجتماع

ش ٢٠٢٤/٣٧/ACD

-اجتمعت لجنة الدراسة بتاريخ ٢٠٢٤/٥/٢٩ لتوضيح الاستفسارات الواردة بمذكرتي شعبة المشتريات المركزية ارقام(ش ٢٠٢٤/٣٧/acd/٢٦٥٧ تاريخ ٢٠٢٤/٥/٢٣)(ش ٢٠٢٤/٣٧/acd/٢٧٣٥ تاريخ ٢٠٢٤/٥/٢٨) وقررت مايلي:

١-بالنسبة لاستفسار السادة شركة سنديان للاتظمة الطبية حول عدد الفحوصات السنوية المطلوبة لفحص (Electrolytes)(ISE) يرجى العلم بانها تعادل عدد فحوصات مادة (CREATINIINE) وهو(2071900)فحص.

٢- مرفق طيه كميات وتوزيع الفحوصات على المواقع المطلوبة فيها.

STOCK #	DESCRIPTION	م.م. الملكة عليا	م.فرج	م.الاميرة ايمان	م.الزرقا	م.الاطفال	م.المسالك	م.الاميرة هيا	م.القلب	م.ماديا	الطوارئ
09_2_000_001	BUN	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_002	CREATININE	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_003	Alb in Urine / CSF	400	0	3000	800	1000	0	800	0	200	0
09_2_000_004	Direct Bilirubin	24000	0	24000	20000	20000	0	15000	2000	9600	18000
09_2_000_005	CPK	29000	7200	18000	36000	15000	0	68000	30000	12000	40000
09_2_000_006	GGT	600	0	6000	1200	1200	0	1200	600	600	0
09_2_000_007	HDL	77000	25000	72000	36000	19000	0	50000	30000	20400	0
09_2_000_008	LDL	77000	25000	72000	36000	19000	0	50000	30000	20400	0
09_2_000_009	Iron	20000	1200	18000	24000	7000	0	24000	2000	4800	0
09_2_000_010	LDH	7000	600	12000	12000	8000	0	15000	15000	600	0
09_2_000_011	Total Protein in Urine / CSF	900	0	7000	2000	2000	0	2400	0	400	0
09_2_000_012	CK- MB	14000	0	0	24000	0	0	20000	17000	4800	30000
09_2_000_014	Mg	4000	0	1200	6000	7000	0	6000	600	0	0
09_2_000_015	Ammonia	0	0	0	1200	4000	0	1000	0	0	0
09_2_000_016	LACTATE	0	0	0	1200	4000	0	1000	0	0	2400
09_2_000_018	Digoxin	0	0	0	0	0	0	0	0	0	0
09_2_000_019	Paracetamol	0	0	0	0	0	0	0	0	0	0
09_2_000_020	Salicylate	0	0	0	0	0	0	0	0	0	0
09_2_000_022	Cholinesterase	0	0	400	0	0	0	0	0	0	0
09_2_000_023	Lipase	0	0	1200	600	0	0	600	0	0	0
09_2_000_024	Apo A lipoprotein	0	0	400	0	0	0	0	0	0	0
09_2_000_025	Apo B lipoprotein	0	0	400	0	0	0	0	0	0	0
09_2_000_026	Hba1c	25000	0	600000	36000	110000	0	20000	12000	15000	0
09_2_000_049	Hba1c/hemolyzing	36	0	857	51	157	0	29	17	21	0
09_2_000_027	Ethanol	0	0	0	1600	0	0	600	0	0	0
09_2_000_028	Triglycerides	77000	25000	72000	36000	19000	0	50000	30000	20400	0
09_2_000_029	Glucose	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_030	Uric Acid	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_031	Cholesterol	77000	25000	72000	36000	19000	0	50000	30000	20400	0
09_2_000_032	Calicum	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000

STOCK #	DESCRIPTION	م.م.الملكة عليا	م.فرج	م.الاميرة ايمن	م.الزرقا	م.الاطفال	م.المسلك	م.الاميرة هيا	م.القلب	م.ماديا	الطوارئ
09_2_000_033	Phosphorus	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_034	ISE DILUENT										
09_2_000_035	Total Bilirubin	80000	25000	120000	70000	90000	0	90000	55000	16800	40000
09_2_000_036	Total Protein	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_037	Alkaline phosphatase	80000	25000	120000	70000	90000	0	90000	55000	16800	40000
09_2_000_038	Albumin	160000	25000	120000	144000	100000	30000	140000	90000	20400	90000
09_2_000_039	ALT	80000	25000	120000	70000	90000	0	90000	55000	16800	40000
09_2_000_040	AST	80000	25000	120000	70000	90000	0	90000	55000	16800	40000
09_2_000_041	Amylase	12000	10000	12000	18000	15000	0	50000	2000	10000	57600

STOCK #	DESCRIPTION	المستعمل	الاختصاص	م.الملك طلال	ع.الخيارجية	م.الامير زيد	م.العقبة	م.الاورام	م.الامير علي	م.الظرون	م.ملركا
09_2_000_001	BUN	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000
09_2_000_002	CREATININE	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000
09_2_000_003	Alb in Urine / CSF	0	600	1200	0	600	300	0	1000	300	0
09_2_000_004	Direct Bilirubin	12000	4800	12000	6000	18000	11000	3600	20000	6000	1000
09_2_000_005	CPK	12000	20000	20000	10000	15000	15000	1200	50000	35000	3000
09_2_000_006	GGT	120	2400	1000	600	1200	1200	240	1200	1200	0
09_2_000_007	HDL	6000	70000	36000	18000	25000	12000	3600	50000	36000	3000
09_2_000_008	LDL	6000	70000	36000	18000	25000	12000	3600	50000	36000	3000
09_2_000_009	Iron	4000	20000	10000	1000	18000	2500	1800	15000	20000	1000
09_2_000_010	LDH	12000	8000	8000	2500	18000	3600	9600	10000	10000	600
09_2_000_011	Total Protein in Urine / C	2000	12000	2000	0	2000	600	0	2000	600	0
09_2_000_012	CK- MB	12000	0	10000	0	6000	6000	0	25000	2000	0
09_2_000_014	Mg	18000	0	1200	0	3000	800	0	1500	2000	0
09_2_000_015	Ammonia	1000	0	600	0	3000	1200	0	2000	0	0
09_2_000_016	LACTATE	1000	0	600	0	3000	1200	0	2000	0	0
09_2_000_018	Digoxin	2500	0	0	0	0	0	0	0	0	0
09_2_000_019	Paracetamol	900	0	0	0	0	0	0	0	0	0
09_2_000_020	Salicylate	600	0	0	0	0	0	0	0	0	0
09_2_000_022	Cholinesterase	0	0	0	0	0	0	0	0	0	0
09_2_000_023	Lipase	0	0	600	0	600	600	0	600	0	0
09_2_000_024	Apo A lipoprotein	0	0	0	0	0	0	0	0	0	0
09_2_000_025	Apo B lipoprotein	0	0	0	0	0	0	0	0	0	0
09_2_000_026	Hba1c	0	36000	20000	18000	10000	12000	0	30000	24000	2000
09_2_000_049	Hba1c/hemolyzing	0	51	29	26	14	17	0	43	34	3
09_2_000_027	Ethanol	1200	0	0	0	300	1000	0	1500	0	0
09_2_000_028	Triglycerides	6000	70000	36000	18000	25000	12000	6000	50000	36000	3000
09_2_000_029	Glucose	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000
09_2_000_030	Uric Acid	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000
09_2_000_031	Cholesterol	6000	70000	36000	18000	25000	12000	6000	50000	36000	3000
09_2_000_032	Calicum	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000

STOCK #	DESCRIPTION	المستعمل	الاختصاص	م.الملك طلال	ع.الخارجية	م.الامير زيد	م.العقبة	م.الاورام	م.الامير علي	م.الظرون	م.ملركا
09_2_000_033	Phosphorus	180000	70000	118000	30000	50000	90000	72000	150000	70000	6000
09_2_000_034	ISE DILUENT										
09_2_000_035	Total Bilirubin	24000	72000	72000	20000	30000	40000	96000	90000	70000	3600
09_2_000_036	Total Protein	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000
09_2_000_037	Alkaline phosphatase	24000	72000	72000	20000	30000	40000	96000	90000	70000	3600
09_2_000_038	Albumin	180000	70000	118000	30000	50000	96000	96000	150000	70000	6000
09_2_000_039	ALT	24000	72000	72000	20000	30000	40000	96000	90000	70000	3600
09_2_000_040	AST	24000	72000	72000	20000	30000	40000	96000	90000	70000	3600
09_2_000_041	Amylase	6000	20000	12000	10000	15000	10000	2400	50000	24000	1000

(4)

STOCK #	DESCRIPTION	م.الامير راشد	القيمة	مجان	الطبية	الشويك	المخبرات	الازرق	اللجان
09_2_000_001	BUN	20000	4000	22000	2000	12000	12000	7500	24000
09_2_000_002	CREATININE	20000	4000	22000	2000	12000	12000	10500	24000
09_2_000_003	ALb in Urine / CSF	600	0	0	0	0	0	0	0
09_2_000_004	Direct Bilirubin	10000	0	2000	0	3000	0	3500	4200
09_2_000_005	CPK	70000	1200	1200	12000	6000	7200	2400	2000
09_2_000_006	GGT	1200	0	0	0	0	0	0	0
09_2_000_007	HDL	40000	3000	22000	15000	6000	9600	0	24000
09_2_000_008	LDL	40000	3000	22000	15000	6000	9600	0	24000
09_2_000_009	Iron	17000	0	1200	4800	4000	2400	0	4300
09_2_000_010	LDH	16000	0	1000	0	4000	3600	0	2400
09_2_000_011	Total Protein in Urine / CSF	2000	0	0	0	0	0	0	0
09_2_000_012	CK- MB	13000	0	0	2000	0	0	0	0
09_2_000_014	Mg	5000	0	0	2400	0	2400	0	0
09_2_000_015	Ammonia	1000	0	0	0	0	0	0	0
09_2_000_016	LACTATE	1000	0	0	0	0	0	0	0
09_2_000_018	Digoxin	0	0	0	0	0	0	0	0
09_2_000_019	Paracetamol	0	0	0	0	0	0	0	0
09_2_000_020	Salicylate	0	0	0	0	0	0	0	0
09_2_000_022	Cholinesterase	0	0	0	0	0	0	0	0
09_2_000_023	Lipase	600	0	0	0	0	0	0	0
09_2_000_024	Apo A lipoprotein	0	0	0	0	0	0	0	0
09_2_000_025	Apo B lipoprotein	0	0	0	0	0	0	0	0
09_2_000_026	Hba1c	35000	2400	4000	4000	20000	6000	4000	22000
09_2_000_049	Hba1c/hemolyzing	50	3	6	6	29	9	6	31
09_2_000_027	Ethanol	2000	0	0	0	0	0	0	0
09_2_000_028	Triglycerides	40000	3000	22000	15000	6000	9600	3750	24000
09_2_000_029	Glucose	200000	4000	22000	2000	12000	12000	12000	24000
09_2_000_030	Uric Acid	200000	4000	22000	2000	12000	12000	6000	24000
09_2_000_031	Cholesterol	40000	3000	22000	15000	6000	9600	6000	24000
09_2_000_032	Calicum	200000	4000	22000	2000	12000	12000	4500	24000

STOCK #	DESCRIPTION	م.الامبر راشد	القوية	معان	الطبية	الشوك	المخبرات	الارزق	التجان
09_2_000_033	Phosphorus	200000	4000	22000	2000	12000	12000	3750	24000
09_2_000_034	ISE DILUENT								
09_2_000_035	Total Bilirubin	170000	3000	22000	12000	6000	12000	3750	24000
09_2_000_036	Total Protein	200000	4000	22000	2000	12000	12000	4500	24000
09_2_000_037	Alkaline phosphatase	170000	3000	22000	12000	6000	12000	6000	24000
09_2_000_038	Albumin	200000	4000	22000	2000	12000	12000	4500	24000
09_2_000_039	ALT	170000	3000	22000	12000	6000	12000	750	24000
09_2_000_040	AST	170000	3000	22000	12000	6000	12000	750	24000
09_2_000_041	Amylase	200000	1800	2400	7200	1000	7200	4500	7000

بسم الله الرحمن الرحيم

محضر اجتماع

ش/٢٠٢٤/٣٧/ACD

عت لجنة الدراسة بتاريخ ٢٠٢٤/٥/١٤ لتوضيح الاستفسارات الواردة بمذكرتي شعبة المشتريات المركزية ارقام (ش/٢٠٢٤/٣٧/٢٤٣٨ تاريخ ٢٠٢٤/٥/١٣) (ش/٢٠٢٤/٣٧/٣٤٥٤ تاريخ ٢٠٢٤/٥/١٣) وقررت مايلي :

برفق طيه مواصفات وعدد وحصة كل جهاز من الاجهزة المطلوبة في العطاء .

المقصود بما ورد في النقطة (٢، ٣) هو فحوصات (NA, K, CL)

لا يوجد مشكلة في احتساب المادة الواردة في النقطة (٢) (Hba1c Hemolyzing) مجانية لانه هناك شركات اخرى بها بسعر.

لا يمكن اعتماد أي فحص دون تجربته في الخدمات الطبية الملكية.

تحذف كلمة REAGENTS من الشرط رقم (١١) الخاص بالمواد المجانية في العطاء.

نعتذر عن تعديل الشرط رقم (١٦)

Item 1

Fully Automated Clinical Chemistry Analyser High Range

Qty. ()

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
1	Brand new Fully automated heavy-duty clinical chemistry analyser floor standing		
2	Discrete and random access processing for routine analysis with stat facility.		
3	Continuous sample loading .no need to stop the system		
4	Accepts sample material of serum, plasma, urine,whole blood ,body fluids and CSF		
5	The system should consist of minimum 50 patient sample positions and stat position should be available and not less than 25 tests available on board for each sample,		
6	The reagent disk consists of reagent position for minimum 25 parameters completely at same time		
7	Ready to use reagents , no need to preparation		
8	No manual pre-treatment of whole blood samples		
9	A minimum throughput of 500 tests/hour without ISE		

	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
10	Sample volume range: 2 - 70 μ l		
11	Reagent volume should be not more than 2 - 400 μ l on board		
12	Stability of reagents up to 3 months.		
13	The instrument should have sample and reagent detection, and automatic dilution.		
14	Computer interface and LCD screen and friendly software.		
15	High quality brand name Laser printer (local supplied printers are accepted)		
16	A list of standard accessories for the offered model		
17	Complete with built in refrigerated compartment (Colling unit not external separate unit).		
18	All the running requirements for the system including reagents, consumables, electrodes , full quality control daily ,calibration and maintenance kits sufficient as annex. (A).		
19	Pre-installation shall be the sole responsibility of the supplier. Pre-installation shall include any civil work, R.O. Water system, electrical work and UPS or site modification(s) necessary to accommodate the new system(s) according to manufacturers' specifications and safety standards in addition to the work required for bringing back the site to the same working conditions as before installing the new systems if required.		
20	Service training for tow(biomedical engineer or biomedical technician) as per term 22.1 of the special terms.		
21	It should support the connection to an internal LAN system or host computer as part of Hospital Information System or Laboratory Information System		
22	Bidirectional communication for Hospital Information System or Laboratory Information System		
23	ISO requirements(reagents, documents, calculation, procedures and reference range)must be included upon request		

Item 2

Fully Automated clinical chemistry Analyser Mid Range

Qty. ()

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
1	Brand new Random Access, fully automated floor standing		
2	A list of standard accessories for the offered model (optional).		
3	A minimum throughput Not less than 300 tests per hour without ISE continuous loading		
4	Continuous sample loading. no need to stop the system		
5	Not less than 25 onboard chemistries		
6	Ready to use reagents , no need to preparation		
7	No manual pre-treatment of whole blood samples		
8	Complete with built-in refrigerated compartment (Colling unit not external separate unit).		
9	Sample and reagent level detection		
10	Wide variety of automatic selection of not less than 10 visible wavelengths: 340-700nm.		
11	Automatic dilution of abnormal readings		
12	Instrument to be attached to intra-net interface system		

	Minimum Requirements	Compliance (Y/N) , Notes	Brochure Page No.
13	Real time of reagent/sample tray reaction status		
14	Stat sample priority not less than 5 samples		
15	High quality brand name Laser printer (local supplied printers are accepted).		
16	Computer interface and LCD screen.		
17	All the running requirements for the system including reagents, consumables, electrodes , full quality control daily ,calibration and maintenance kits sufficient as annex. (A).		
18	Pre-installation shall be the sole responsibility of the supplier. Pre-installation shall include any civil work, R.O. Water system, electrical work and UPS or site modification(s) necessary to accommodate the new system(s) according to manufacturers' specifications and safety standards in addition to the work required for bringing back the site to the same working conditions as before installing the new systems if required.		
19	Service training for tow(biomedical engineer or biomedical technician) as per term 22.1 of the special terms.		
20	It should support the connection to an internal LAN system or host computer as part of Hospital Information System or Laboratory Information System		
21	Bidirectional communication for Hospital Information System or Laboratory Information System		
22	ISO requirements(reagents, documents, calculation, procedures and reference range)must be included upon request		
23	Automatic probe cleaning		

١. لا يتم الإحالة على شركة أجنبية ما لم يكن لها وكيل أردني أو مكتب مسجل في الأردن حسب الأصول
٢. يلتزم الوكيل بالاتي:
 - أ- تقديم أسعار المواد التشغيلية (reagents, consumables and disposables) لعمل كافة الفحوصات المطلوبة والمبينة في الملحق (A) ملتزم بها
 - ب-
 - ت- يجب أن تكون جميع المواد التشغيلية (reagents, consumables and disposables) من نفس الشركة الصانعة للأجهزة المقدمة وتكون من نفس المنشأ ويشمل ذلك التصنيع و التعبئة.
 - ث-
 - ج- بيان تاريخ الانتهاء وظروف التخزين على جميع المواد.
 - ح- يجب على الوكيل استبدال اي مادة يثبت عدم جودتها عند الاستعمال.
 - خ- يحق لمديرية الخدمات الطبية الملكية في حال ظهور مشاكل في جودة المواد المستخدمة ودقة نتائجها وتوثيق ذلك من قبل مديرية الخدمات الطبية الملكية ان تقوم بإلغاء التزامها بشراء المواد من الوكيل وذلك دون ان يترتب عليها اية تكاليف اضافية.
 - د-
 - ذ- يحق لمديرية الخدمات الطبية الملكية زيادة كميات المواد أو الاجهزه المحالة في حال تطلب الأمر وعلى الوكيل تقديمها بنفس الأسعار والشروط المبينة في جداول قرار الإحالة وبنسبه لا تتجاوز ٣٠%.
 - ر- يحق لمديرية الخدمات الطبية تأجيل التوريد أو إلغاء او استبدال أي كمية من الكواشف أو المواد على أن يتم إبلاغ الوكيل بذلك قبل شهرين من موعد التوريد .
 - ز- تؤخذ المواد والكواشف لكل جهاز كوحدة واحدة من نفس الشركة الصانعة.
 - س- تلتزم الشركة بتقديم شهادات FDA or CE لجميع الفحوصات المطلوبة .
٣. يلتزم الوكيل بصيانة الأجهزة شاملة قطع الغيار للأجهزة لمدة طول مدة العطاء مجاناً.
٤. يلتزم الوكيل بان يشمل الإصلاح على أعمال الصيانة الفنية والكشف على كفاءة الجهاز وإجراء الفحوصات اللازمة لذلك وأجراء الصيانة الدورية الوقائية وإصلاح الأعطال وتقديم قطع الغيار وتركيبها بما فيها المستهلكات مثل الالكترودات ومواد المعايرة بالإعداد والكميات اللازمة لتشغيل الجهاز على مدار الساعة وطيلة مدة الاتفاقية وكلما تطلب ذلك دون ان تتحمل الخدمات الطبية الملكية اية نفقات مالية إضافية.
٥. يعتبر جزءاً من الجهاز كل ما توصي به الشركة الصانعة كضرورة لعمل الجهاز بشكل طبيعي حتى وان لم يرد ذلك ضمن المواصفات الفنية الواردة في العطاء.
٦. تعتبر جميع الأجهزة ملكاً للخدمات الطبية الملكية.
٧. يحق للوكيل المحلي احتساب الاجهزه الموجودة في الخدمة في الخدمات الطبية الملكية من عدد الأجهزة المطلوبة في الملحق A بعد موافقة الخدمات الطبية الملكية وحسب الشروط التالية :
 - أ- أن لا يكون قد تجاوز العمر التشغيلي لهذه الاجهزه عن ست سنوات ل (High Range and MidRange) من تاريخ التشغيل الى تاريخ إغلاق العطاء.

- ب- في حال تقدمت الشركة بأجهزة جديدة فيجب ان تتطابق مع مواصفات الاجهزة المرفقة والشروط الخاصة وتعليمات الخدمات الطبية في الاجهزة عند الشراء وحسب مستوى الجهاز
- ت- أن تخضع هذه الأجهزة لجميع شروط العطاء
- ث- ان لا يكون هناك اي مشاكل او ملاحظات على هذه الاجهزة
- ج- ان تتحمل الشركة نقل وتشغيل اي جهاز تم اعتماده الى اي موقع تراه الخدمات الطبية الملكية مناسباً
- ح- ان تكون الاجهزة المحتسبة من نفس الموديلات للاجهزة المناقص عليها وان تكون لها نفس المواد التشغيلية ولا تحتاج الى اي مواد اضافية
- خ- يتم المناقصة على كامل المحاليل والمستهلكات لجميع عدد الاجهزة المطلوبة سواء كانت هذه الاجهزة جديدة او في الخدمة .
٨. يجب ان تقوم الشركة بتقديم كامل الاجهزة من نفس الصنف ولايجوز تقديمها بشكل جزئي، وفي حال عدم تطابق الاجهزة المطلوبة من الصنف الواحد مع الكمية المطلوبة يعتبر العرض مخالفاً.
٩. يحق للشركة ان تتقدم بصنف اعلى من الصنف المطلوب وذلك بعد موافقة الخدمات الطبية الملكية بما يتلائم مع طبيعة العمل والموقع.
١٠. يحق للشركة المناقصة على جميع فئات الاجهزة او أي فئة معينة من الاجهزة بشرط ان تكون المناقصة على كل فئة بشكل كامل.
١١. يتم التعامل مع المحاليل المطلوبة لكل موقع على اساس وحدة واحدة سواء كانت الاحالة بشكل كامل او منوع .
١٢. يحق للخدمات الطبية الملكية الاحالة بشكل كامل او جزئي وحسب ماتراه اللجنة مناسباً.
١٣. تقسم المواد بالتساوي بين عدد الاجهزة في الموقع الذي يوجد فيه اكثر من جهاز من فئة واحدة فقط
١٤. توزع المواد حسب عدد الاجهزة واعتماداً على الاسعار بالنسب التالية وبشكل تقريبي

عدد اجهزه high	عدد اجهزه mid	نسبة المواد على high	نسبة المواد على mid
1	1	%60	%40
1	2	%40	30%,30%
2	2	30%, 30%	20%, 20%
2	3	25%25%	50%/3

SPECIAL TERMS

- Offers not complying with any of the special terms or the technical specifications will be considered non-conforming.
 - Any vendor providing FORGED documents will be disqualified from the current tender and any future RMS tenders.
1. All equipment must be the most recently released model/version which is equal to or higher than the range of the specifications of the required system (low, mid or high) and equal to or higher than the level of technology and requested options mentioned in the technical specifications.
 2. Required certificates:
 - a. For equipment of US origin, a copy of a certificate of FDA approval & the relevant 510K clearance for selling to US healthcare facilities for the offered model must be submitted with the technical offer.
 - b. For equipment of other origins, a copy of either a CE certificate (MDD)/TÜV OR a certificate of FDA approval & the relevant 510K clearance for selling to US healthcare facilities for the offered model must be submitted with the technical offer.
 - c. Only for class I medical equipment, submission of a copy of Declaration of Conformity certificate (MDD) for the offered model is accepted.
 - d. For the following types of medical equipment a certificate of FDA approval & the relevant 510K clearance for selling to US healthcare facilities must be provided (ventilators, infant incubators, anaesthesia machines, heart lung machines, balloon pumps, defibrillators, AED machines, patient monitors, resuscitators, auto-pulse devices, vessel sealing system).
 - e. In all of the above cases certificates must be formally endorsed by JFDA.
 3. Country of origin:
 - a. The country of origin of the main part (s) of the system must be:
USA, Canada, Japan, UK, Sweden, Finland, Denmark, Switzerland, Belgium, Germany, France, Netherlands, Spain, Norway, Italy, Ireland, Austria, New Zealand and Australia.

- b. Sub-components of the system, accessories and consumables (as determined by the purchasing committee) are accepted to be manufactured in other countries & by different manufacturers.*
 - c. All offered items must be approved for sale in the same country of origin.*
 - d. Vendors must specify the origin of all offered items in the technical offer.*
 - e. Equipment manufactured by reputable companies based in any of the countries mentioned in (3.A) will be taken into consideration regardless of the country of origin only if they are approved for sale in at least three of the countries mentioned in (3.A), an original and officially endorsed free-sale certificate must be included in the offer.*
- 4. Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first third of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period will be rejected.*
- 5. Warranty:*
- a. Offers must include a full warranty including spare parts and labour for a period of a minimum of 24 months from the date of installation or 30 months from the date of receiving the items at the agreed location mentioned in the final order.*
 - b. In the case where a delay in installation has occurred as a result of the supplier's dereliction and has exceeded a period of one month from the date of receiving the items, the approved warranty period will automatically be considered as 24 months from the date of installation.*
 - c. If at any time during the warranty period the item becomes inoperative due to a technical fault the item must then be repaired by the supplier /local agent within a period of thirty days from written notification, otherwise the supplier must replace the item with a new functioning one and will endure a penalty determined by the Royal Medical Services for each day of the downtime of the system. In case the item was replaced by a new one, the warranty period mentioned in term No.(5-A) above will start from the installation and commissioning date of the new item.*
- 6. One set of operation manual(s) and one set of service manual(s) including schematics and a spare-part list must be delivered with each unit, CD/DVD is acceptable.*
- For large tenders, a certain agreed percentage of manuals per each item will be accepted.*

7. *Where applicable, pre-installation shall be the sole responsibility of the supplier. Pre-installation shall include removal of the old system, any civil work, electrical work or site modifications necessary to accommodate the new system according to manufacturers' specifications and safety standards in addition to the work required for bringing back the site to the same original condition before installing the new system.*
8. *Power requirements: where applicable either single phase 220V, 50Hz or 3-phase 380V. Systems with external transformers are only considered conforming only if stated in the technical specifications.*
9. *Technical offers must include clear original technical brochures/catalogues for all offered items.*
10. *Offers must include fully detailed technical offers and compliance sheets as a soft copy (either Microsoft office or Microsoft excel format) in addition to a hard copy format, mentioning the exact model/catalogue number and country of origin of the offered item(s), full technical description/specifications and any accessories or options included in the offer.*
11. *Compliance sheets must be as per the format provided with the tender documents for the technical specifications of the tender listing the required specification on one side and a Yes or NO response to each point on the other side with reference to page and line numbers in the specific technical brochure.*

Offers not complying with this term will be rejected.

12. *Any accessories and consumable items necessary to operate the offered system must be clearly identified and priced.*
13. *Technical offers must include a priced list for accessories and consumables as a hard copy in addition to a soft copy format copy (either Microsoft office or Microsoft excel) with prices fixed for a period of five years commencing on the date of the start of the warranty period.*

Prices must be fixed with a maximum annual increase of 5%.

Accessories and consumables must be priced according to their delivery destination either to QueenAliaInternationalAirport or to RMS Main Medical Stores.

14. *Technical offers must include a comprehensive and priced spare parts list as a hard copy in addition to a soft copy format copy (either Microsoft office or Microsoft excel) valid for a minimum period of five years commencing at the end date of the warranty period.*

Prices must be fixed with a maximum annual increase of 5%.

Spare parts must be priced according to their delivery destination either to QueenAliaInternationalAirport or to RMS Main Medical Stores.

15. Where applicable, a start-up kit of accessories and consumable items must be provided with each system on a free of charge basis.

16. Spare parts, consumables and accessories availability must be guaranteed for a minimum period of ten years starting from date of installation and commissioning.

17. For the final list of offers having equal chances of winning the award, the awarding process will be based on the accumulative value of both the offered item and its' running cost over a period of seven years.

Running cost includes the value of consumables, accessories & service contract price (where applicable) needed to operate the system over the same period.

Only offers with the lowest sum value (system price and its running cost) will qualify for the award.

18. a. For PC/Laptop based systems; a complete restoration CD/DVD/etc (operating system and application software) must be supplied.

b. Where locally supplied computers or laptops are offered, only computers / laptops from Apple, hp/Compaq, Lenovo, Dell, Fujitsu&Toshiba will be accepted, offered models must be the latest available version upon delivery.

c. Where locally supplied printers are offered only the following types and brands are accepted:

LASER: HP, SAMSUNG, OKI, CANON

DOT MATRIX: EPSON, OKI.

19. Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores, installation, pre-installation (if needed), training, commissioning, warranty and bringing the equipment into service.

20. Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).

21. DRMS has the right to increase or decrease the awarded quantities by a percentage not exceeding 30% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.

22. **Training:**

22.1 For items where abroad service training courses for the offered system are usually conducted, offers must include a certified service training program at a reputable center abroad recognized by the manufacturer for at least one biomedical engineer or biomedical technician; all costs inclusive, air tickets, , boarding, commuting, accommodation (minimum 3 star hotel on full board basis) and any extra costs.

22.2 For items where abroad user training courses for the offered item are usually conducted offers must include a certified operator training program at a reputable center abroad recognized by the manufacturer for at least one operator; all costs inclusive, air tickets, , boarding, commuting, accommodation (minimum 3 star hotel on full board basis) and any extra costs.

22.3 The period of the training courses must be according to the manufacturer's program and must be stated clearly in the technical offer.

22.4 Programs must conform to the following standards:

- User training must comprise understanding and use of operation manual(s), correct and safe operation of the equipment, and user preventive maintenance and calibration.
- Service training must comprise: theory, understanding and use of service manual(s), calibration, preventive maintenance procedure, and practical troubleshooting and repair exercises, and must be conducted by professional instructors employed or authorized by the system manufacturer.
- Service training must be conducted on a system of identical make, model, and configuration to that purchased by DRMS, and designated by the manufacturer or the local agent for training purposes.
- Certificates must be endorsed and officially sealed by the system manufacturer, legally empowering trainees to engage in user and service activities according to operation and service manual(s).

22.5 For all items and where applicable, offers must include an on-site user and service training.

Annex (a)

الموقع	عدد اجهزة high	عدد اجهزة medium
الملكة عليا	2	2
مركز فرح	-	1
المركز	4	-
الزرقاء	2	2
الاطفال	2	-
المسالك	-	1
الاميرة هيا	1	2
القلب	1	2
مادبا	-	2
الطوارئ	2	-
المستعمل	2	1
الاختصاص	2	-
الملكطلال	1	2
الخارجية	1	1
الامير زيد	1	2
العقبة	1	2
مركز الاورام		
الامير علي	1	3
اللطرون	2	2
ماركا	-	1
الامير راشد	3	2
القوية	-	1
معان	-	1
الطباية	-	1
الشوبك	-	1
المخابرات	-	1
الازرق	-	1
اللجان	-	1

DIRECTORATE ROYAL MEDICAL SERVICES

**reagents and cosumables for automated chemistry
department**

TENDER NO P37-2024-ACD

09_2 Biochemistry Department

09_2_000_Reagent For AUTOMATED CHEMISTRY DEPARTMENT

1	09_2_000_001	2068900	TEST	BUN
2	09_2_000_002	2071900	TEST	CREATININE
3	09_2_000_003	10800	TEST	ALb in Urine / CSF
4	09_2_000_004	249700	TEST	Direct Bilirubin
5	09_2_000_005	538400	TEST	CPK
6	09_2_000_006	21760	TEST	GGT
7	09_2_000_007	708600	TEST	HDL
8	09_2_000_008	708600	TEST	LDL
9	09_2_000_009	228000	TEST	Iron
10	09_2_000_010	179500	TEST	LDH
11	09_2_000_011	37900	TEST	Total Protein in Urine / CSF
12	09_2_000_012	185800	TEST	CK- MB
13	09_2_000_014	61100	TEST	Mg
14	09_2_000_015	15000	TEST	Ammonia
15	09_2_000_016	17400	TEST	LACTATE
16	09_2_000_018	2500	TEST	Digoxin
17	09_2_000_019	900	TEST	Paracetamol
18	09_2_000_020	600	TEST	Salicylate
19	09_2_000_022	400	TEST	Cholinesterase
20	09_2_000_023	5400	TEST	Lipase
21	09_2_000_024	400	TEST	Apo A lipoprotein
22	09_2_000_025	400	TEST	Apo B lipoprotein
23	09_2_000_026	1067400	TEST	Hbalc
24	09_2_000_027	8200	TEST	Ethanol
25	09_2_000_028	714750	TEST	Triglycerides
26	09_2_000_029	2073400	TEST	Glucose

DIRECTORATE ROYAL MEDICAL SERVICES

**reagents and cosumables for automated chemistry
department**

TENDER NO P37-2024-ACD

09_2 Biochemistry Department

27	09_2_000_030	2067400	TEST	Uric Acid
28	09_2_000_031	717000	TEST	Cholesterol
29	09_2_000_032	2065900	TEST	Calicum
30	09_2_000_033	2035150	TEST	Phosphorus
31	09_2_000_034	492	PACK	ISE Diluent
32	09_2_000_035	1357150	TEST	Total Bilirubin
33	09_2_000_036	2065900	TEST	Total Protein
34	09_2_000_037	1359400	TEST	Alkaline phosphatase
35	09_2_000_038	2065900	TEST	Albumin
36	09_2_000_039	1354150	TEST	ALT
37	09_2_000_040	1354150	TEST	AST
38	09_2_000_041	388100	TEST	Amylase
39	09_2_000_049	1525	TEST	HbA1c Hemolysing reagent (45ML/KIT) ,For Cobas C501
40	09_2_000_050	764	PACK	ISE INTERNAL STANDARD
41	09_2_000_051	292	PACK	ISE REFERENCE

Handwritten signature

الشروط الخاصة للعطاء رقم reagents and cosumables for automated P37-2024-ACD

chemistry department

الشروط الخاصة التسلسل

- 1 - Number of shipments will be decided by purchasing committee according to the respective minimum guaranteed shelf life upon delivery/dispatch to be provided in the original offers to establish delivery schedule accordingly. The DRMS with this respect reserves the right to reject any item not in compliance with this term and impose the correspondent fines as per GHQ/Military Procurement Laws
- 2 - Dispatched/Delivered goods must show Description of Goods, Shelf Life, Expiry Date, Batch No., Storage conditions, and any other necessary information.
- 3 - Goods must be dispatched/delivered under the same storage conditions that comply with their nature.
- 4 - Offer to include all kit size configuration from Manufacturer.
- 5 - All Packages to be stamped or labelled with the following" Sold to RMS" Tender P 37-2024-ACD"
- 6 - Any Products derived or containing materials of human origin should be supplied with a certificate and/or official document from the manufacturer stating that the products are devoid of blood borne infectious agents: HbsAg, HIV, HCV, etc . Items belonging to this category to be specified in the original offer.
- 7 - The quality of items to be Previously accepted and approved by DRMS according to DRMS established regulations.
- 8 - Number of tests per Kit/Pack to be clearly mentioned where applicable.

الشروط الخاصة للعطاء رقم P37-2024-ACD reagents and cosumables for automated

chemistry department

الشروط الخاصة التسلسل

- 9 - The DRMS reserves the right to purchase goods in respect to their kit size (tests/kit) which will correlate with the technical needs of the specific department.
- 10 - When the DRMS decides to purchase any consumables and reagents for the equipments , anew free maintenance warranty will be effective immediatly,including spare parts and labour.
- 11 - Any Additional running requirements for the system including reagents,consumables,electrodes,full quality control daily,calibration,lamps,reaction cells sets,sample probe substrate,reagent probe,sample cups,sample cuvettes,cuvette probe wash,cleaning solutions,maintenance kits and any other necessary requirements for the system must be mentioned clearly and to be supplied in sufficient amounts as free of charge according to the number of tests
- 12 - Acopy of (CE OR FDA certificated)is to be submitted and should be Certificated By JFDA (jordan food & drug administration)

الشروط الخاصة للعطاء رقم P37-2024-ACD reagents and cosumables for automated
chemistry department

الشروط الخاصة التسلسل

- 13 - Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces(JAF), however , suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the ports corporation or QAIA (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges ,local shipping charges etc.) The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s)
- 14 - Pricing must include services of sale, shipment , transportation ,delivery from port to site or Main Medical Stores.
- 15 - For offers submitted in Jordanian dinars payment will be either by wire transfer or by cheque to be paid after receving goods ,Any other way of payment will not be accepted and the offers will be rejected by the purchase committee.
- 16 - All items should retain at least (8) mounths for MSL .Any items not complying with this term upon receipt should be accompanied with confirmation to replace any remaining unused quantities after the expiry date .Said confirmation is subject to the approval of the director of Royal Medical Services and will incur afine, which will be decided later by DRMS according to the loss that this discrepancy with term of the tender has caused.

الشروط الخاصة للعطاء رقم P37-2024-ACD reagents and cosumables for automated

chemistry department

الشروط الخاصة التسلسل

17 - Bidders must submit their reservation /queries regarding tender specifications and /or special term within the first half of the tender closing period starting from the tender announcement date .

Reservations/queries submitted after the end of this period shall be rejected.

18 - Adequte on site user training and service training must be provided.

19 - Acomprehensive priced spare parts list valid for 5 years after the full warranty period should be submitted.

20 - Spare parts must be available for minimum period of 10 years from date of insallation.