

مديرية الخدمات الطبية الملكية



نظام المشتريات والعطاءات

اسم العطاء : P37-2024-DRY  
وصف العطاء : CHEMISTRY  
Reagents For Dry  
Chemistry  
تاريخ إنشاء العطاء: 2024/10/13  
الصلاحية:  
الوثيقة المالية:  
طريقة الشراء:

Item No	Stock No	Stock Name	Quantity	Unit
1	09_2_0000_020	General Chemistry Tests Disc: (Na,K,Ca,Urea,Creatinine,UA,Phos,Alb ,TP,TBIL,ALT,AST,ALP,Glu,CK,TG,Ch olesterol)	19600	Disc
2	09_2_0000_021	Chemistry Controls Set(2-levels)For Automatic Dry Chemistry Analyzer	132	Kit

---

---

### شروط العطاء

---

### شروط خاصة

- 1-Storage condition as well as description of good ,batch NO. and expiry date must be clearly indicated on all packages and relative document.
- 2-- goods must be dispatched/delivered under the same storage conditions that comply with their nature.
- 3-All packages to be stamped or labelled with the following "Sold to RMS" Tender P37-2024-Dry Chemistry"
- 4-Goods should be previously evaluated and approved in DRMS or purchased by "central procurement branch " or by " directorate of defence procurement ".
- 5- Products derived or containing materials of human origin should be supplied with a certificate and/or official document from the maunfacturer stating that the products are devoid of blood borne infectious agents :HbsAg, HIV, HCV, ect. Items belonging to this category to be specified in the original offer.
- 6-ALL itemes must have FDA,TUV,OR CE Approval.
- 7-Number of test per pack to be clearly mentioned.
- 8-For offers submitted in Jordanian dinars payment will be either by wire transfer or by cheque to be paid after reciving documents from main medical stores ,any other way of payment will not be accepted and the offers will be rejected by the purchase committee.
- 9-Samples and catalogues of the quoted items should be submitted along with the offers for purchasing and receipt purposes showing the same serial number/cat as per

---

our tender documents specifications. RMS with this reserves the right to reject offers not in compliance with this terms.

10-The supplier or his local agent must furnish RMS a guarantee stamped and legalized by the notary Public to 115% of the goods valid for 1 year after Acceptance of good.

11--Custom clearance of goods shall be the responsibility of the Jordanian armed forces(JAF)however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the ports corporation or qala (including expenses caused by delay in presenting the necessary document for either clearing or transporting the goods to the required location mentioned in the final order ,delivery note issuing charges, unloading charges, local shipping charges etc).the supplier is also responsible for providing of all relevant shipping documents ,together with the delivery order(s).

12--Pricing must include services of sale, shipment, transportation, and delivery From port to site or to main medical stores.

13- Number of shipments will be decided by purchasing committee according to the respective minimum guaranteed shelf life upon delivery/dispatch to be provided in the original offers to establish delivery schedule accordingly .The DRMS with this respect reserves the right to reject any item not in compliance with this term and impose the correspondent fines.

14-The supplier must keep the existing machine up running to the expected performance level.

15-Bidders must submit their reservation /queries regarding tender specificareqions and/or special term within the first half of the tender closing period starting from the tender announcement date . Reservation /queries submitted after the end of this period shall be rejected.

16-ALL additional requirements for test performance should be supplied as a free of charge according to each shipment separately sufficient to the hospital such as (cups, cuvett,control,calibrators,pobes,lamps wash solution,resin ...etc.)

17- All goods should retain at least the MSL 9 months. Any item not complying with this term upon receipt should be accompanied with a confirmation that you accept to replace any remaining unused quantities at your expenses after the expiry date. Said confirmation is subject to the approval of the director of the Royal Medical Services and will incur a fine, which will be decided later by DRMS according to the loss that this discrepancy with terms of the tender has caused.

تكفل الطرف الآخر بتزويد الخدمات الطبية الماكينة باجهزة -18

عدد (11) (Full Automatic Dry Chemistry Analyzer)

على سبيل الإعارة إلى قيادة المستودعات الطبية الرئيسية مقابل شراء المستهلكات وتكون ملكيتها للشركة وتعاد عند إنتهاء الكميات التي تم إحالتها على الشركة بموجب العطاء (العطاءات اللاحقة) وتكون الشركة مسؤولة عن تشغيل وتركيب الأجهزة وصيانتها وإدامة عملها والتدريب عليها لكافة العاملين على الجهاز وفي جميع اماكن تواجد الجهاز خلال فترة التشغيل وفي حال تعطل أحد الأجهزة على الشركة صيانة الجهاز خلال مدة لا تتجاوز 48 ساعة من التبغ بالعطل وإذا زادت مدة التعطل عن هذه المدة تكون ملزمة بتقديم جهاز بديل لإدامة العمل خلال 48 ساعة على أن يتم تحديد موعد تسليم الجهاز من قبل لجنة الشراء

يجب على الوكيل استبدال اي مادة تثبت عدم فعاليتها وجودتها عند الاستعمال -19