



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 1	Wheel Chair, Adult	Qty. (175)
--------	--------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Adult size wheel-chair (seat width 42-44 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 2	Wheel Chair, Extra Wide Adult	Qty. (50)
--------	-------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24. inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Extra wide Adult size wheel-chair (seat width 56-60 cm)		



JORDAN ARMED FORCES

The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 3	Wheel Chair, Children	Qty. (50)
--------	-----------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Children size wheel-chair (seat width 36-38 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 4	Wheel Chair, Wide Adult	Qty. (175)
--------	-------------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	wide Adult size wheel-chair (seat width 46-48 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 5	Wheel Chair, Narrow Adult	Qty. (40)
--------	---------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Narrow Adult size wheel-chair (seat width 38-40cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 6	WHEELCHAIR, WIDE ADULT(FULLY RECLINING BACK)	Qty - (30)
--------	--	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair wide Adult size (seat width 46-48 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 7	WHEELCHAIR, ADULT(FULLY RECLINING BACK)	Qty. (30)
--------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Adult size (seat width 42-44 cm)		



Item 8	WHEELCHAIR,NARRW ADULT(FULLY RECLINING BACK)	Qty. (10)
--------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Narrow Adult size (seat width 38-40 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 9	WHEELCHAIR, CHILDREN (FULLY RECLINING BACK)	Qty. (40)
--------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be Pneumatic (Pump & English key are Essential) or rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair children size (seat width 36-38cm)		

Item 10	WHEELCHAIR FOR CHILDREN WITH CEREBRAL PALSY(36cm)	Qty. (25)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction aluminium frame.		
2	Fully reclining back (90-180) degree with head support wheel-chair (seat width 36 cm)		
3	Adjustable seat angle.		
4	Detachable desk arms rest		
5	Pneumatic tired wheels for outdoor use (pump & English key are essential)		
6	Frame made for durable, anti-corrosion, easy to clean		
7	Durable front caster solid or pneumatic to be 360 pivot type, with locking mechanism		
8	Adjustable height head rest (for head stability)		
9	Adjustable height foot plates		
10	Padded abductor for thigh separation		
11	Padding material should be durable, provides optimum comfort, easy to clean and disinfect.		
12	The wheelchair should have foldable seat and backrest.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 11	WHEELCHAIR FOR CHILDREN WITH CEREBRAL PALSY(30cm)	Qty. (60)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction aluminium frame.		
2	Fully reclining back (90-180) degree with head support wheel-chair (seat width 30 cm)		
3	Adjustable seat angle.		
4	Detachable desk arms rest		
5	Pneumatic tired with wheels for outdoor use (pump & English key are essential)		
6	Frame made for durable, anti-corrosion, easy to clean		
7	Durable front caster solid or pneumatic to be 360 pivot type, with locking mechanism		
8	Adjustable height head rest (for head stability)		
9	Adjustable height foot plates		
10	Padded abductor for thigh separation		
11	Padding material should be durable, provides optimum comfort, easy to clean and disinfect.		
12	The wheelchair should have foldable seat and backrest.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 12	Commode chair wide adult size (seat 46-48 cm)	Qty. (75)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details			
Name of Manufacturer			
Model/ catalogue number			
Country of Origin for the offered model			
Country where the manufacturer is based			
Delivery time			
Full warranty period			
FDA clearance OR CE Mark			

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Wide Adult size self-propelled shower & commode chair (seat width 46-48 cm).		
2	Flip up arm rest.		
3	Weight capacity \geq 135 kg		
4	Pneumatic tired nylon wheels with hand rims with(pump & English) or rigid type		
5	Solid front castors wheels with lock		
6	Detachable / swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 13	Commode chair adult size (seat 42-44cm)	Qty. (75)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size self-propelled shower & commode chair (seat width 42-44 cm).		
2	Flip up arm rest.		
3	Weight capacity ≥ 100 kg		
4	Pneumatic tired nylon wheels with hand rims with(pump & English) or rigid type		
5	Solid front castors wheels with lock		
6	Detachable / swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 14	Walker, wide adult	Qty. (75)
---------	--------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Wide adult size walker. Approximately 95 cm height 55cm width		
2	Adjustable height. 30cm – 40cm		
3	Aluminium bar.		
4	Foam hand grip.		



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 15	Walker, adult	Qty. (75)
---------	---------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Normal adult size walker Approximately 95 cm height 48cm width		
2	Adjustable height. 30cm – 40cm		
3	Aluminium bar.		
4	Foam hand grip.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 16	K-Walker, large	Qty. (25)
---------	-----------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 70cm – 80cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 17	K-Walker, small	Qty. (35)
---------	-----------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 2 to 7.		
3	Adjustable height. 55cm – 65cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 18	Two fronts Walker style children size (2-7)	Qty. (20)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 2 to 7.		
3	Adjustable height. 45cm – 55cm		
4	Aluminium bars.		
5	Padded hand grip.		
6	Two front wheels		



JORDAN ARMED FORCES

The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUE OF BIOMEDICAL TECHNOLOGY



Item 19	Standing Frame style (supine/prone) medium size	Qty. (30)
---------	---	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size standing frame.		
2	For children 7-15 years.		
3	Manual height adjustment of top.		
4	lift up to around 140kgs patient from sitting to standing position.		
5	Dimensions of top please specify with raised rim.		
6	Cut-out size is approximately		
7	Padded knee.		
8	Chest and back supports holds patient in standing position.		
9	Stainless steel frame or equivalent		
10	Retractable locking swivel casters.		
11	Frame base is please specify.		
12	Padded Head support .		
13	Straps: high quality vinyl nylon with velecro 10cm width, Qty(4).		
14	Reattachable abduction pillow.		



JORDAN ARMED FORCES
The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
15	Incline degrees (85-60-30-0)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 20	Wheelchair pure gel medical cushion (43*41*3 cm)	Qty. (50)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm)		
3	Cushion for adult size wheel-chair (43cm width * 41 length)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 21	Wheelchair pure gel medical cushion (40*40*3 cm)	Qty. (20)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm.)		
3	Cushion for narrow adult size wheel-chair (40cm width * 40 cm length).		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 22	Wheelchair pure gel medical cushion (36*28*3 cm)	Qty. (10)
---------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion or equivalent		
2	Height (3 cm.)		
3	Cushion for children size wheel-chair (36cm width * 28cm length).		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 23	Elbow crutches (Tall adult)	Qty. (10)
---------	-------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Tall Adult size stick.130cm – 150cm height		
2	Strong, lightweight aluminium crutch.		
3	Adjustable height.		
4	Padding hand grip		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 24	Elbow crutches (Adult)	Qty. (30)
---------	--------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size stick. 110 – 130 cm height		
2	Strong, lightweight aluminium crutch.		
3	Adjustable height.		
4	Padding hand grip		



JORDAN ARMED FORCES
The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 25	K-Walker, medium	Qty. (30)
---------	------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 60 -70 cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 26	Standing Frame style (supine/prone) small size	Qty. (80)
----------------	--	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size standing frame.		
2	For children 2-7years.		
3	Manual height adjustment of top .		
4	Unit can lift up to around 140kgs patient from sitting to standing position.		
5	Padded head support .		
6	Dimensions of top with raised rim please specify all dimensions.		
7	Cut-out size is approximately		
8	Padded knee		
9	Chest and back supports holds patient in standing position.		
10	Stainless steel frame or equivalent		
11	Retractable locking swivel casters.		
12	Frame base please specify		
13	Straps: high quality vinyl nylon with velcro 10cm width, Qty(4).		
14	Removable abduction pillow.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
15	Incline degrees (85-60-30-0)		



SPECIAL TERMS

- *Offers not complying with any of the special terms or the technical specifications shall be considered non-conforming with tender requirements.*
 - *Any vendor providing FORGED documents shall be disqualified from the current tender and banned from participating in any future RMS tenders.*
 - *Wherever term "based" is mentioned it refers to the country where the manufacturing company is founded & established.*
1. *All equipment must be the most recently released model/version which is equal to or higher than the range of the specifications of the required system (low, mid or high) and equal to or higher than the level of technology and required options mentioned in the technical specifications.*
 2. *Required certificates (must be submitted with the technical offer):*
 - 2.1 *FDA clearance (510K) for equipment of USA based companies.*
 - 2.2 *MDR (EU) 2017/745 certificate for equipment of EU (European Union) based companies.*
 - 2.3 *UKCA certificate For Equipment of Great Britain based companies (England, Scotland, Northern Ireland, and Wales).*
 - 2.4 *ARTG (Australian Register of Therapeutic Goods) certification or approval for Australian and New Zealand based companies.*
 - 2.5 *Establishment License with the relevant Device License issued by the Therapeutic Products Directorate for Equipment of Canadian based companies.*
 - 2.6 *PMDA (Pharmaceuticals and Medical Devices Agency) certification or approval for equipment of Japanese based companies.*
 - 2.7 *Swissmedic (Swiss agency for therapeutic products) certification or approval for equipment of Swiss based companies.*
 - 2.8 *Norwegian Medicines Agency certification or approval for equipment of Norwegian based companies.*
 - 2.9 *Only for class I medical equipment manufactured by companies based in one of the countries mentioned above, submission of either one of the certificates mentioned above or a free sale certificate in any of these countries shall be accepted.*

[Handwritten signature]



JORDAN ARMED FORCES

The DIRECTION OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



2.10 For the following equipment:

- All X-ray equipment, MRI, ultrasound, and nuclear medicine systems (regardless where the manufacturing companies are based).
- Where the manufacturing companies are based in other origins than the mentioned in terms 2.1 - 2.8.

The following are required:

- a- At least two of the certificates mentioned above, one of which has to be FDA clearance (510K) (Only for class I medical equipment submission of certificates mentioned in 2.9 shall be accepted).
- b- Evaluation certificate from the Royal Medical Services for the same offered model with at least 80% passing grade.

If the evaluation is not applicable (based on purchasing committee perspective) bidder should submit a list of installation basis of the same offered model and/or previous models in at least two of the following hospitals (King Hussein Cancer Center, National Center for Diabetes Endocrinology and Genetic Diseases, Jordan University Hospital or King Abdullah University Hospital) with at least three years of operation, list should include: Name of hospital, Model installed, Quantity, and date of installation.

The purchasing committee has the right to officially contact any of these hospitals and disqualify any offer where the feedback is negative in operation, after sales service or local agent performance.

2.11 The vendor is responsible to ensure through official documents that classified medical devices are manufactured in conformity with applicable quality system standards (ISO, IEC); (the international quality systems standard for medical device manufacturing ISO 13485) or ISO9001.

2.12 With each offer, bidders must provide a formally endorsed document issued by the manufacturer stating that the bidder is the sole certified agent for the offered item.

2.13 In all of the above cases (except 2.12) certificates must be formally endorsed by JFDA.

2.14 Any vendor not submitting all required certificates will be eliminated.

3. Official website of the manufacturer must be provided. The website must clearly demonstrate the history profile and manufacturing features of the company including the same offered model and its brochure.

In case of inconsistency between the information on the website with the provided information of the local agent regarding any offered product, the committee has the right to eliminate that product from the awarding process.



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



4. Offered items should be from reputable well known manufacturers and excellent experience in the field and shall have multiple installations of the same offered model and/ or previous models in RMS main hospitals with at least two years of operation and excellent experience in operation, after sales service & local agent performance; otherwise the purchasing committee has the right to request any of the following:

- a) An evaluation certificate as mentioned in term 2.10.b.
- b) A sample of the same offered model at any time during the purchasing process, the sample unit shall be delivered within three weeks from the date of a written notification; any vendor rejects to deliver a sample unit will be eliminated. Evaluation duration will be determined by the purchasing committee.

Any offered item fail in the evaluation/assessment process will be rejected

5. Vendors must specify the origin of the offered items and accessories in the technical offer.
6. Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first half of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period shall be rejected.
7. Goods are to be brand new; manufacturing date of the delivered products should not exceed 18 months from the date of the final award.
8. Delivery period should be mentioned clearly; the purchasing committee has the right to reject any offer with delivery time exceeding 4 months.

9. Warranty:

- a- Offers must include a full warranty for a period of a minimum of 24 months from the date of installation or 30 months from the date of receiving the items at the agreed location mentioned in the final order (whichever comes first).

Warranty must include corrective & preventive maintenance activities as per manufacturer recommendations including:

- Required spare parts (free of charge)
- Labour
- Hardware
- Software
- Rechargeable batteries

At the end of the warranty period, The supplier commits to implement final inspection of the submitted goods and submit the reports signed by the site chief engineer stating that the equipment are working properly as well as all preventive maintenance reports during the warranty period.



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Warranty & installation are excluded for items mentioned in Attachment no.1 the purchasing committee has the right to exclude any other item not listed in the attachment and does not required installation and warranty.

- b- In the case where a delay in installation has occurred as a result of the supplier's dereliction and has exceeded a period of one month from the date of receiving the items, the approved warranty period will automatically be considered as 24 months from the date of installation.
 - c- If at any time during the warranty period the item becomes inoperative due to a technical fault the item must then be repaired by the supplier /local agent within a period of 5 working days from written notification, warranty will be extended according to downtime period.
 - d- If the delay exceeds 30 days the supplier must replace the item with a new identical functioning one (within the same delivery period mentioned in the final offer). In case the item was replaced by a new one, the warranty period mentioned in 9.a) above will start from the installation and commissioning date of the new item.
 - e- Local agent/ supplier is committed to transport and install any awarded item during the warranty period on free of charge basis to any location inside the country whenever required by Royal Medical Services, this should not include pre-installation requirements in the new site.
10. All offered items (main unit) should be fully designed, manufactured, and labelled by their real original manufacturer in which all related testing, research, development and approvals went through.
- Any relabelled products for the main unit (white-label manufacturing, OEM, or repackaging) are rejected.
11. One set of operation manual(s) and one set of service manual(s) including schematics and a spare-part list must be delivered with each unit, CD/DVD is acceptable. For large tenders, a certain agreed percentage of manuals per item may be agreed upon.
12. Where applicable, pre-installation shall be the sole responsibility of the supplier. Pre-installation shall include removal of old system(s), any civil work, electrical work or site modification(s) necessary to accommodate the new system(s) according to manufacturers' specifications and safety standards in addition to the work required for bringing back the site to the same working conditions as before installing the new system(s).



13. *Power requirements: where applicable either single phase 220V, 50Hz or 3-phase 380V. Systems with external transformers are considered conforming only if clearly stated in the technical specifications.*
14. *Technical offers must include clear original technical brochures/catalogues for all offered items.*
15. *Offers must include fully detailed technical offers and compliance sheets as a soft copy (either Microsoft office or Microsoft excel format) in addition to a hard copy, mentioning the exact model/catalogue number and country of origin of the offered item(s), full technical description/specifications and any accessories or options included in the offer.*
Offer must clearly indicate the origin of the offered item and the country where the manufacturer of the offered items is based.
16. *Compliance sheets must be as per the tabular format of the technical specifications in the tender documents, listing the required specifications on one column and a Yes or NO response to each point in the adjacent column, with reference to page and line numbers in the relevant technical brochure. Offers not complying with this term shall be rejected.*
17. *Qualifications and after sales service: The technical bid should contain all the necessary documents (training certificates) to prove the capability of the bidders for supplying and installing a trouble free equipment meeting the quality standards and technical specifications of the manufacturer and the ability of the bidders for providing efficient after sales service conducted by authorized certified biomedical service engineers with minimum 2 years experience in the same field.*
18. *Accessories and consumables:*
- 18.1 *Any accessories and consumable items necessary to operate the offered system must be clearly identified and priced separately.*
All offered accessories and consumables must be approved by the manufacturer.
- 18.2 *Technical offers must include a priced list for all accessories and consumables related to the required equipment as a hard copy in addition to a soft copy (either Microsoft office or Microsoft excel format) with prices fixed for a period of five years from the date of installation and commissioning with a maximum annual increase of 2%, any essential item not listed will be considered free of charge.*
- 18.3 *Accessories and consumables must be priced according to their delivery destination either to Queen Alia International Airport or to RMS Main Medical Stores.*
- 18.4 *Where applicable, a start-up kit of accessories and consumable items must be provided with each system on a free-of-charge basis.*



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



19. Spare Parts:

- 19.1 *Technical offers must include a comprehensive and priced list of all spare parts related to the awarded equipment (including rechargeable batteries) as a hard copy in addition to a soft copy (either Microsoft office or Microsoft excel format) valid for a minimum period of five years with a maximum annual increase of 2%, commencing at the end date of the warranty period, any essential item not listed will be considered free of charge.*
 - 19.2 *Spare parts must be priced according to their delivery destination either to Queen Alia International Airport or to RMS Main Medical Stores.*
 - 19.3 *Prices of spare parts should be reasonable and will be taken into consideration during the purchasing process; the purchasing committee has the right to eliminate any offered item with unreasonable high prices of spare parts.*
 - 19.4 *Delivery period of required spare parts should not exceed 2 months from the date of the final order.*
20. *Spare parts, consumables and accessories availability must be guaranteed for a minimum period of ten years starting from the date of installation and commissioning.*
21. *Local agent is committed to sign a service contract with Royal Medical Services to perform all relevant maintenance for the awarded equipment (whenever required by DRMS within a period that does not exceed one year from the end date of the warranty period and if mentioned clearly in the specifications) without exceeding the original mentioned cost and as per mentioned terms within minimum required period.*

22. Tender Awards:

- 22.1 *For the final list of offers having a chance of winning the award, the awarding process shall be based on the accumulative value of both the offered item and its' running cost (Total Cost of Ownership) over a period of seven years from the date of installation and commissioning. Only offers with the lowest total cost of ownership over a period of seven years from the date of installation and commissioning shall qualify for the award.*
- 22.2 *Running cost includes the value of consumables, accessories needed to operate the system over the same period as well as the cost of any service contract (where applicable).*



23. For PC/Laptop based systems:

- 23.1 Complete restoration medium (CD/DVD/etc.) of the operating system and the application software must be supplied.**
- 23.2 Purchased licenses, software keys & dongles must be provided at any time during and after the warranty period on free of charge basis whenever required by DRMS for a minimum period of ten years starting from the date of installation and commissioning.**
- 23.3 Where locally supplied computers, laptops & printers are offered, the offered model should be from well known manufacturer.**

24. Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores, installation, pre-installation (if needed), training, commissioning, warranty and bringing the equipment into service.

25. Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).

26. a. DRMS has the right to increase the awarded quantities by a percentage not exceeding 35% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.

b. DRMS has the right to decrease the awarded quantities by a percentage not exceeding 50% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.

27. a. Bidder is not allowed to submit more than one offer for the same item whether that's solely or in coalition or partnership with other bidders.

b. Bidder is allowed to include within their offer optional alternatives for the same offered item from the same manufacturer.

28. The supplier must furnish DRMS with a guarantee stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded equipment valid for twelve months from the date of final acceptance of the equipment by DRMS.

011



29. Training:

- 29.1 *For items where abroad service training courses are required in technical specifications, offers must include a certified service training program for at least 3 working days at a reputable center abroad recognized by the manufacturer for at least one biomedical engineer or biomedical technician; all costs inclusive, air tickets, , boarding, commuting, accommodation (minimum 3 star hotel on full board basis) and any extra costs.*
- 29.2 *For items where abroad user training courses are required in technical specifications, offers must include a certified operator training program at a reputable center abroad recognized by the manufacturer for at least one operator; all costs inclusive, air tickets, boarding, commuting, accommodation (minimum 3 star hotel on full board basis) and any extra costs.*
- 29.3 *The period of the training courses must be according to the manufacturer's program excluding traveling days and must be stated clearly in the technical offer.*
- 29.4 *Training Programs must conform to the following standards:*
- *User training must comprise understanding and use of operation manual(s), correct and safe operation of the equipment, as well as user preventive maintenance and calibration.*
 - *Service training must comprise: theory, understanding and use of service manual(s), calibration, preventive maintenance procedure, and practical troubleshooting and repair exercises, and must be conducted by professional instructors employed or authorized by the system manufacturer.*
 - *Service training must be conducted on a system of identical make, model, and configuration to that purchased by DRMS, and designated by the manufacturer or the local agent for training purposes.*
 - *Certificates must be endorsed and officially sealed by the system manufacturer, legally empowering trainees to engage in user and service activities according to operation and service manual(s).*
 - *Where applicable, offers must include an on-site user and service training.*

30. For offers submitted in Jordanian dinars , payment will be either by wire transfer or by cheque after final acceptance of goods .Any other way of payment will be rejected .

31. Samples:

A .Bidders must submit samples with the offer for purchasing & receiving purpose .

B .Receiving of goods will be compared with samples accepted by purchasing committee.



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



• Attachment 1

#	Equipment
1	Adjustable Auto Pipettes
2	Resuscitation Bag
3	Laryngoscope Set
4	Oxygen Flow meter wall type/ single
5	Regulator Suction with canister, wall vacuum outlet
6	Oxygen Regulator for Oxygen Cylinders
7	Pulse Oximeter, Finger type
8	Oxygen Cylinder
9	Doppler, portable
10	Diagnostic set, Portable
11	Direct Ophthalmoscope, Portable
12	Otoscope, Portable
13	Air Mattress System, homecare
14	Stethoscope
15	Aneroid Sphygmomanometer
16	Video Assisted Laryngoscope, portable
17	Scale, Manual
18	Wood's Light
19	Cough Pressure, Normal Saline
20	Rehabilitation Walking Parallel Bars, non-powered
21	Therapy Mat
22	Medical Ball All Size
23	Dumbbells Rack with complete set of dumbbells
24	Crutches
25	Shoulder wheel
26	Mobile Mirror
27	Cuff Weights
28	Walker, different sizes
29	Patient Elbow Stick
30	Rehabilitation Training Ladder
31	Rehabilitation Suspension Frame
32	Exercise Band All Size (Theraband)



JORDAN ARMED FORCES
The DIRCETROATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



33	<i>Lens trial set</i>
34	<i>Wheel Chair</i>
35	<i>Commode Chair</i>
36	<i>Bassinet (Baby Cot)</i>
37	<i>Resuscitation Cart (Crash Cart)</i>
38	<i>Medication Cart</i>
39	<i>Cart, Drawers</i>
40	<i>Examination Couch, Manual</i>
41	<i>Gynaecology Examination Table, Manual</i>
42	<i>Examination Table, Neonates, Manual</i>
43	<i>Intravenous Pole, Mobile Stand</i>
44	<i>DDA Cabinet</i>
45	<i>Stainless steel Multipurpose Trolley</i>
46	<i>Cabinet, Instrument, Operation Theatres</i>
47	<i>Dressing Cart</i>
48	<i>Stainless steel wire shelving unit</i>
49	<i>Paper Trolley</i>
50	<i>Stainless Steel Sink (Clean up counter)</i>
51	<i>Scopes Cabinet</i>
52	<i>Mayo Table</i>
53	<i>Table, Instrument</i>
54	<i>Stool, Adjustable, Doctor</i>
55	<i>Stool, Adjustable, Operation Theatres</i>
56	<i>Carts, linen/laundry, soiled, Double</i>
57	<i>Closed distribution trolley</i>
58	<i>Step Ladder, Conductive, Double</i>
59	<i>Step, Surgeon, Single</i>
60	<i>Kick Bucket</i>
61	<i>Mobile Stand for Oxygen Cylinder</i>
62	<i>Stainless Steel Wire Basket, 1 STU</i>
63	<i>Cart, Plaster</i>