



### محضر اجتماع

تموضع : العطاء رقم شن ٢٧/٢٠٢٤/٥٠٠ لشراء جهاز Laproscopic Surgical Unit

اجتمعت لجنة مواصفات العطاء اعلاه و المشتملة بموجب كتاب عطوفة مدير عام الخدمات الطبية الملكية رقم شن ٤٤١٢/٢٧/٢٠٢٤/٥٠٠ بتاريخ ٢٠٢٤/٧/٢٦

استفسل السادة شركة المهتمون للخدمات والتوريدات الطبية الاردنية رقم تاريخ 774/2024 ٢٠٢٤/٨/٢٢

تعمل المواصفة رقم (A-1) لتصبح :

بلا من :

CCD or CMOS camera head

At least 3 chip CCD or CMOS camera head

• لا تعديل على باقي المواصفات والشروط الخاصة



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 1	Laparoscopic Surgical Unit	Qty. (2)
--------	----------------------------	----------

**IMPORTANT NOTE:**

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

**TECHNICAL SPECIFICATIONS:**

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
Durable, heavy duty construction, reliable with superior image quality: the committee has the right to request for a sample of the same offered model at any time during the purchasing process, the sample unit shall be delivered within three weeks from the date of a written notification, any vendor rejects to deliver a sample unit will be eliminated		
<b>A- Camera system (4k Resolution) : Provides electronic shutter function, automatic brightness control and optical zoom</b>		
1) At least 3 chip CCD or CMOS camera head		
2) Resolution (4k): a minimum of 3840*2160.		
3) Compact, lightweight, easy to grip with low heat generation		
4) Focal Distance with zoom lens, 16-28 mm or better		
5) Gas sterilizable, autoclavable is preferable		
6) Immersible in disinfections liquid.		
7) Universal coupling mechanism to fit all standard endoscopes.		
8) Camera cable 3 m or more.		
9) 4k Camera controller.		
10) Picture in picture mode.		






# JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
<b>B- Light Source:</b>		
1) LED Light source (light intensity should be equivalent to 300 Watt xenon lamp)		
2) Working hour $\geq$ 10,000 hours.		
3) Automatic and manual adjustment of light intensity.		
4) Front panel hygienic with convenient indicators for ease of operation.		
5) Flexible fibre optic cable preferably autoclavable and not less than 230 mm Provides optimal light transmission and long durability		
6) Only Fibre optic type ( Led lamp inside the console) accepted if the light source is integrated with camera console		
7) LED module should be priced separately.		
<b>C- CO2 Insufflator:</b>		
1) Sealed hygienic front touch panel.		
2) Maximum flow rate From 40- 50 L/MIN.		
3) Display of gas consumption, flow rate and intra-abdominal pressure		
4) Gas heating to body temperature (at 37 C).		
5) Audible and visual alarms.		
6) Filter on the CO2 output to prevent cross-contamination.		
7) Smoke evacuator that is either integrated or quoted as a separate unit.		
<b>D- Suction/ Irrigation combined unit:</b>		
1) Roller pump for irrigation (Pressure: adjustable up to 200 mmHg or better).		
2) Flow Rate: adjustable up to 1 litre/min or better.		
3) Adjustable vacuum pressure.		
4) Complete with all standard sets of tubes for the unit.		
<b>E- LCD / LED Colour Monitor: Qty. (1)</b>		
1) Medical grade type that is supplied by the same manufacturer		





# JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
2) At least of 32 inch size that is mounted on the mobile cart		
3) Outstanding image quality with 4k resolution minimum of 3840*2160		
4) DVI compatible.		
<b>F- Mobile cart:</b> Original to suit the above systems and consists of the following:		
1) 4 antistatic, anti-scratch castors. And four shelves		
2) Locking brakes.		
3) Cable channel with shockproof plugs.		
4) Adjustable or fixed shelves.		
5) Articulating swivel arm mount for the monitor		
6) At least one drawer.		
7) All accessories (connect cable, plugs, etc.) needed for the offered system to be included.		
<b>G- Ability to record images and videos on Medical grade DVD recorder or USB flash memory</b>		
<b>H- Telescopes, 4K Resolution :</b>		
1) TELESCOPE 30° Diameter 5 mm		
2) TELESCOPE 0° Diameter 5 mm		
3) TELESCOPE 30° Diameter 10 mm		
4) TELESCOPE 0° Diameter 10 mm		

01



# JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
<p><b>I- Upgrading of one system to be capable of Near Infrared Fluorescence Imaging (ICG application) with the below features should be priced separately:</b></p> <ul style="list-style-type: none"><li>a- White light image with superimposed display of NIR/ICG fluorescence.</li><li>b- Possible to select the preferred color for NIR/ICG imaging: Either blue or green.</li><li>c- NIR/ICG fluorescence signal in white. Background in black for maximum contrast.</li><li>d- White light image with superimposed display of NIR/ICG fluorescence. NIR/ICG signal display will appear in different colors depending on the strength of the detected NIR signal.</li><li>e- with all necessary hardware/accessories/consumables</li></ul>		
<p><b>J- Any essential hardware/accessories/consumables required to make all above items functional should be quoted, otherwise it will be supplied free of charge. (In case of only disposable type of items the cost of 100 case will be added to the price of the system)</b></p>		
<p><b>K- Abroad service training for 1 biomedical engineers/ technicians as per special terms</b></p>		



## JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



### SPECIAL TERMS

- *Offers not complying with any of the special terms or the technical specifications shall be considered non-conforming with tender requirements.*
- *Any vendor providing FORGED documents shall be disqualified from the current tender and banned from participating in any future RMS tenders.*
  1. *All equipment must be the most recently released model/version which is equal to or higher than the range of the specifications of the required system (low, mid or high) and equal to or higher than the level of technology and required options mentioned in the technical specifications.*
  2. *Required certificates (must be submitted with the technical offer):*
    - 2.1 *FDA approval & the relevant 510K clearance for equipment of USA origin (Country where the manufacturer is based).*
    - 2.2 *CE certificate with the relevant CE number (Medical directive) for equipment of EU (European Union) origin (Country where the manufacturer is based).*
    - 2.3 *UKCA certificate For Equipment of Great Britain origin (England, Scotland, Northern Ireland, and Wales) (Country where the manufacturer is based).*
    - 2.4 *ENTID (Enterprise Identification) with the relevant ARTG (Australian Register of Therapeutic Goods) number issued by the Therapeutic Goods Administration for Equipment of Australian origin (Country where the manufacturer is based).*
    - 2.5 *Establishment License with the relevant Device License issued by the Therapeutic Products Directorate for Equipment of Canadian origin (Country where the manufacturer is based).*
    - 2.6 *Japanese Pharmaceutical Affairs Law (JPAL) certification or approval for equipment of Japanese origin (Country where the manufacturer is based).*
    - 2.7 *Swissmedic (Swiss agency for therapeutic products) certification or approval for equipment of Swiss origin (Country where the manufacturer is based).*
    - 2.8 *Norwegian Medicines Agency certification or approval for equipment of Norwegian origin (Country where the manufacturer is based).*
    - 2.9 *The vendor is responsible to ensure through official documents that classified medical devices are manufactured in conformity with applicable quality system standards (ISO, IEC); (the international quality systems standard for medical device manufacturing ISO 13485) or ISO9001.*



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



4. *Country of origin:*

4.1. *The country of origin of the main part (s) of the system must be one of the following:*

*USA, Canada, Japan, UK, Sweden, Finland, Denmark, Switzerland, Belgium, Germany, France, Netherlands, Spain, Norway, Italy, Ireland, Austria, New Zealand, Australia, Czech Republic, Luxembourg & Poland.*

4.2. *Accessories and consumables may be manufactured in other countries and/or by different manufacturers and should be approved by the manufacturer.*

4.3. *All offered items must be approved for sale in the same country of origin. An original and officially endorsed free-sale certificate from an authorised body must be included in the offer.*

4.4. *Vendors must specify the origin of all offered items and accessories in the technical offer.*

4.5. *Except for X-ray based equipment, MRI, and nuclear medicine systems, equipment manufactured by reputable companies based in any of the countries mentioned in (4.1) will be taken into consideration regardless of the manufacturing site only:*

*a. If they are approved for sale in the same county of origin (an original and officially endorsed free-sale certificate from an authorised body must be included in the offer.*

*OR*

*b. If they are approved for sale in at least three of the countries mentioned in (4.1) (an original and officially endorsed free-sale certificate from an authorised body in those countries must be included in the offer).*

5. *Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first half of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period shall be rejected.*

6. *Goods are to be brand new; manufacturing date of the delivered products should not exceed 18 months from the date of the final award.*

7. *Delivery period should be mentioned clearly; the purchasing committee has the right to reject any offer with delivery time exceeding 4 months.*

8. *Warranty:*

*a- Offers must include a full warranty for a period of a minimum of 24 months from the date of installation or 30 months from the date of receiving the items at the agreed location mentioned in the final order (whichever comes first).*



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



*Warranty must include corrective & preventive maintenance activities as per manufacturer recommendations including:*

- *Required spare parts (free of charge)*
- *Labour*
- *Hardware*
- *Software*
- *Rechargeable batteries*

*At the end of the warranty period, The supplier commits to implement final inspection of the submitted goods and submit the reports signed by the site chief engineer stating that the equipment are working properly as well as all preventive maintenance reports during the warranty period.*

*Warranty & installation are excluded for items mentioned in Attachment no.1 the purchasing committee has the right to exclude any other item not listed in the attachment and does not required installation and warranty.*

- b- In the case where a delay in installation has occurred as a result of the supplier's dereliction and has exceeded a period of one month from the date of receiving the items, the approved warranty period will automatically be considered as 24 months from the date of installation.*
  - c- If at any time during the warranty period the item becomes inoperative due to a technical fault the item must then be repaired by the supplier /local agent within a period of 5 working days from written notification, warranty will be extended according to downtime period.*
  - d- If the delay exceeds 30 days the supplier must replace the item with a new identical functioning one (within the same delivery period mentioned in the final offer). In case the item was replaced by a new one, the warranty period mentioned in 8.a) above will start from the installation and commissioning date of the new item.*
  - e- Local agent/ supplier is committed to transport and install any awarded item during the warranty period on free of charge basis to any location inside the country whenever required by Royal Medical Services, this should not include pre-installation requirements in the new site.*
9. *One set of operation manual(s) and one set of service manual(s) including schematics and a spare-part list must be delivered with each unit, CD/DVD is acceptable. For large tenders, a certain agreed percentage of manuals per item may be agreed upon.*





JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



10. *Where applicable, pre-installation shall be the sole responsibility of the supplier. Pre-installation shall include removal of old system(s), any civil work, electrical work or site modification(s) necessary to accommodate the new system(s) according to manufacturers' specifications and safety standards in addition to the work required for bringing back the site to the same working conditions as before installing the new system(s).*
11. *Power requirements: where applicable either single phase 220V, 50Hz or 3-phase 380V. Systems with external transformers are considered conforming only if clearly stated in the technical specifications.*
12. *Technical offers must include clear original technical brochures/catalogues for all offered items.*
13. *Offers must include fully detailed technical offers and compliance sheets as a soft copy ( either Microsoft office or Microsoft excel format) in addition to a hard copy, mentioning the exact model/catalogue number and country of origin of the offered item(s), full technical description/specifications and any accessories or options included in the offer.  
Offer must clearly indicate the origin of the offered item and the country where the manufacturer of the offered items is based*
14. *Compliance sheets must be as per the tabular format of the technical specifications in the tender documents, listing the required specifications on one column and a Yes or NO response to each point in the adjacent column, with reference to page and line numbers in the relevant technical brochure. Offers not complying with this term shall be rejected.*
15. *Qualifications and after sales service: The technical bid should contain all the necessary documents (training certificates) to prove the capability of the bidders for supplying and installing a trouble free equipment meeting the quality standards and technical specifications of the manufacturer and the ability of the bidders for providing efficient after sales service conducted by authorized certified biomedical service engineers with minimum 2 years experience in the same field.*
16. *Accessories and consumables:*
  - 16.1 *Any accessories and consumable items necessary to operate the offered system must be clearly identified and priced separately.  
All offered accessories and consumables must be approved by the manufacturer.*
  - 16.2 *Technical offers must include a priced list for all accessories and consumables related to the required equipment as a hard copy in addition to a soft copy (either Microsoft office or Microsoft excel format) with prices fixed for a period of five years from the date of*



## JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



*installation and commissioning with a maximum annual increase of 2%, any essential item not listed will be considered free of charge.*

16.3 *Accessories and consumables must be priced according to their delivery destination either to Queen Alia International Airport or to RMS Main Medical Stores.*

16.4 *Where applicable, a start-up kit of accessories and consumable items must be provided with each system on a free-of-charge basis.*

### 17. Spare Parts:

17.1. *Technical offers must include a comprehensive and priced list of all spare parts related to the awarded equipment (including rechargeable batteries) as a hard copy in addition to a soft copy (either Microsoft office or Microsoft excel format) valid for a minimum period of five years with a maximum annual increase of 2%, commencing at the end date of the warranty period, any essential item not listed will be considered free of charge.*

17.2. *Spare parts must be priced according to their delivery destination either to Queen Alia International Airport or to RMS Main Medical Stores.*

17.3. *Prices of spare parts should be reasonable and will be taken into consideration during the purchasing process; the purchasing committee has the right to eliminate any offered item with unreasonable high prices of spare parts.*

17.4. *Delivery period of required spare parts should not exceed 2 months from the date of the final order.*

18. *Spare parts, consumables and accessories availability must be guaranteed for a minimum period of ten years starting from the date of installation and commissioning.*

19. *Local agent is committed to sign a service contract with Royal Medical Services to perform all relevant maintenance for the awarded equipment (whenever required by DRMS within a period that does not exceed one year from the end date of the warranty period and if mentioned clearly in the specifications) without exceeding the original mentioned cost and as per mentioned terms within minimum required period.*

### 20. Tender Awards:

20.1. *For the final list of offers having a chance of winning the award, the awarding process shall be based on the accumulative value of both the offered item and its' running cost (Total Cost of Ownership) over a period of seven years from the date of installation and commissioning. Only offers with the lowest total cost of ownership over a period of*



JORDAN ARMED FORCES

The DIRCETROATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



*seven years from the date of installation and commissioning shall qualify for the award.*

- 20.2. *Running cost includes the value of consumables, accessories needed to operate the system over the same period as well as the cost of any service contract (where applicable).*
21. *For PC/Laptop based systems:*
- 21.1. *Complete restoration medium (CD/DVD/etc.) of the operating system and the application software must be supplied.*
- 21.2. *Purchased licenses, software keys & dongles must be provided at any time during and after the warranty period on free of charge basis whenever required by DRMS for a minimum period of ten years starting from the date of installation and commissioning.*
- 21.3. *Where locally supplied computers or laptops are offered, only computers/laptops from Apple, hp/Compaq, Lenovo, Dell, Fujitsu or Toshiba will be accepted, offered models must be the latest available version upon delivery.*
- 21.4. *Where locally supplied printers are offered only the following types and brands are accepted: HP, SAMSUNG, OKI, CANON, EPSON.*
22. *Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores, installation, pre-installation (if needed), training, commissioning, warranty and bringing the equipment into service.*
23. *Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).*
24. a. *DRMS has the right to increase the awarded quantities by a percentage not exceeding 35% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.*
- b. *DRMS has the right to decrease the awarded quantities by a percentage not exceeding 50% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.*



JORDAN ARMED FORCES


The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



25. a. Bidder is not allowed to submit more than one offer for the same item whether that's solely or in coalition or partnership with other bidders.
- b. Bidder is allowed to include within their offer optional alternatives for the same offered item from the same manufacturer.
26. The supplier must furnish DRMS with a guarantee stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded equipment valid for twelve months from the date of final acceptance of the equipment by DRMS.
27. Training:
- 27.1 For items where abroad service training courses are required in technical specifications, offers must include a certified service training program at a reputable center abroad recognized by the manufacturer for at least one biomedical engineer or biomedical technician; all costs inclusive, air tickets, , boarding, commuting, accommodation (minimum 3 star hotel on full board basis) and any extra costs.
- 27.2 For items where abroad user training courses are required in technical specifications, offers must include a certified operator training program at a reputable center abroad recognized by the manufacturer for at least one operator; all costs inclusive, air tickets, boarding, commuting, accommodation (minimum 3 star hotel on full board basis) and any extra costs.
- 27.3 The period of the training courses must be according to the manufacturer's program excluding traveling days and must be stated clearly in the technical offer.
- 27.4 Training Programs must conform to the following standards:
- User training must comprise understanding and use of operation manual(s), correct and safe operation of the equipment, as well as user preventive maintenance and calibration.
  - Service training must comprise: theory, understanding and use of service manual(s), calibration, preventive maintenance procedure, and practical troubleshooting and repair exercises, and must be conducted by professional instructors employed or authorized by the system manufacturer.
  - Service training must be conducted on a system of identical make, model, and configuration to that purchased by DRMS, and designated by the manufacturer or the local agent for training purposes.
  - Certificates must be endorsed and officially sealed by the system manufacturer, legally empowering trainees to engage in user and service activities according to operation and service manual(s).
  - Where applicable, offers must include an on-site user and service training.
- [Handwritten signature]*



JORDAN ARMED FORCES  
THE DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



28 For offers submitted in Jordanian dinars, Payment will be either by wire transfer or by cheque after final acceptance of goods. Any other way of payment will be rejected



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



*Attachment 1*

#	Equipment
1	<i>Adjustable Auto Pipettes</i>
2	<i>Resuscitation Bag</i>
3	<i>Laryngoscope Set</i>
4	<i>Oxygen Flow meter wall type/ single</i>
5	<i>Regulator Suction with canister, wall vacuum outlet</i>
6	<i>Oxygen Regulator for Oxygen Cylinders</i>
7	<i>Pulse Oximeter, Finger type</i>
8	<i>Oxygen Cylinder</i>
9	<i>Doppler, portable</i>
10	<i>Diagnostic set, Portable</i>
11	<i>Direct Ophthalmoscope, Portable</i>
12	<i>Otoscope, Portable</i>
13	<i>Air Mattress System, homecare</i>
14	<i>Stethoscope</i>
15	<i>Aneroid Sphygmomanometer</i>
16	<i>Video Assisted Laryngoscope, portable</i>
17	<i>Scale, Manual</i>
18	<i>Wood's Light</i>
19	<i>Cough Pressure, Normal Saline</i>
20	<i>Rehabilitation Walking Parallel Bars, non-powered</i>
21	<i>Therapy Mat</i>
22	<i>Medical Ball All Size</i>
23	<i>Dumbbells Rack with complete set of dumbbells</i>
24	<i>Crutches</i>
25	<i>Shoulder wheel</i>
26	<i>Mobile Mirror</i>
27	<i>Cuff Weights</i>
28	<i>Walker, different sizes</i>
29	<i>Patient Elbow Stick</i>
30	<i>Rehabilitation Training Ladder</i>
31	<i>Rehabilitation Suspension Frame</i>
32	<i>Exercise Band All Size (Theraband)</i>



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES  
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



33	<i>Lens trial set</i>
34	<i>Wheel Chair</i>
35	<i>Commode Chair</i>
36	<i>Bassinet (Baby Cot)</i>
37	<i>Resuscitation Cart (Crash Cart)</i>
38	<i>Medication Cart</i>
39	<i>Cart, Drawers</i>
40	<i>Examination Couch, Manual</i>
41	<i>Gynaecology Examination Table, Manual</i>
42	<i>Examination Table, Neonates, Manual</i>
43	<i>Intravenous Pole, Mobile Stand</i>
44	<i>DDA Cabinet</i>
45	<i>Stainless steel Multipurpose Trolley</i>
46	<i>Cabinet, Instrument, Operation Theatres</i>
47	<i>Dressing Cart</i>
48	<i>Stainless steel wire shelving unit</i>
49	<i>Paper Trolley</i>
50	<i>Stainless Steel Sink (Clean up counter)</i>
51	<i>Scopes Cabinet</i>
52	<i>Mayo Table</i>
53	<i>Table, Instrument</i>
54	<i>Stool, Adjustable, Doctor</i>
55	<i>Stool, Adjustable, Operation Theatres</i>
56	<i>Carts, linen/laundry, soiled, Double</i>
57	<i>Closed distribution trolley</i>
58	<i>Step Ladder, Conductive, Double</i>
59	<i>Step, Surgeon, Single</i>
60	<i>Kick Bucket</i>
61	<i>Mobile Stand for Oxygen Cylinder</i>
62	<i>Stainless Steel Wire Basket, 1 STU</i>
63	<i>Cart, Plaster</i>