



نظام المشتريات والعطاءات

2025/04/28

تاريخ إنشاء العطاء:

P1-2025-5

اسم العطاء:

الصلاحية:

Drapes for Cardiac Surgery

وصف العطاء:

الوثيقة المالية:

طريقة الشراء:

Item No	Stock No	Stock Name	Quantity	Unit
1	20_09_500	Cardiac Surgery Pack I (CABAG: Coronary Artery Bypass Graft) 1. The fabric used in gowns and drapes should abide by one of those standards: 1.1. European Standards CE EN 13795 which include the following technical test specifications: 1.1.1Barrier: a.Resistance to dry microbial penetration(EN ISO22612) b.Resistance to wet microbial penetration(EN ISO 22610) c.Resistance to liquid penetration(EN ISO 20811) 1.1.2. Cleanliness of fabric before sterilization(EN ISO11737-) 1.1.3 Linting(ISO11737-1) 1.1.4 Strength:#a.Bursting strength-Dry and wet(EN 13938-1)#b.Tensile strength-Dry and wet(EN29073-3) 1.1.5 Non inflammable. 1.2. American Standards FDA (ANSI/AAMI PB 70,2012 Level4) 2. One cardiac surgery incise drape (250-270cm)x(180-200cm)with the following specifications: a.Armed covers, b.chest adhesive incise area (28-33cm)x(35-40cm), c.Leg adhesive incise area(50-60cm)x(90-105cm), with reinforced fabric bridge(15-25cm)in	2000	Each

## Drapes for Cardiac Surgery

Item No	Stock No	Stock Name	Quantity	Unit
		<p>between the two windows(for both points b&amp;c), d.Adhesive drape used in window b&amp;c showed, e.Fabric reinforcement, f.Heavy duty fabrics,g.longitudinal flaps with reinforcement, h. Perfusion tubing pockets with clipping tapes(plastic) i.Tubeholders "built-in".</p> <p>3. Two reinforced surgical gowns(XXL); full reinforcement of gown inside anteriorly and both sides entirely whole length WITH TWO ABSORBANT TOWELS.</p> <p>4. Two reinforced surgical gowns(XL); full reinforcement of entire gown inside anteriorly and both sides whole length WITH two ABSORBANT TOWELS.</p> <p>5. One back table cover ,X-large, reinforced180cmx260cm(175-185)(255-270).</p> <p>6. One Mayo stand cover,reinforced 60cmx135cm(55-65)(130-140).</p> <p>7. Two OR cloth towels blue.</p> <p>8. One diathermy bag 35cmx40cm.</p> <p>9. One Bilateral Under-leg drape with perineal cover and umbilicus cover (210x210)cm</p>		
2	20_09_501	<p>Cardio Vascular Pack II: Congenital &amp; valve.</p> <p>1.The fabric used in gowns and drapes standards:</p> <p>1.1. European Standards CE EN13795 which include the following technical tests:</p> <p>1.1.1Barrier: a.Resistence to dry microbial penetration(EN ISO22612), b.Resistence to wet microbial penetration (EN ISO22610), c.Resistence to liquid penetration (EN ISO 20811),</p> <p>1.1.2 Cleanliness of fabric before sterilization (EN ISO11737-1)</p> <p>1.1.3 Linting(ISO 11737-1)</p>	1500	Each

## Drapes for Cardiac Surgery

Item No	Stock No	Stock Name	Quantity	Unit
		1.1.4 Strength: a.Bursting strength-Dry and wet(EN13938-1), b.Tensile strength-Dry and wet(EN 29073-3)		
		1.1.5 Non inflammable.		
		1.2. American Standards FDA (ANSI/AAMI PB70,2012Level4)		
		2. One back_table coverXL reinforced (175-185cm)x(255-270cm)		
		3. One mayo stand cover reinforced(55-65 cm)X(130-140CM)		
		4. Two reinforced surgical gowns(XXL) with entire full reinforcement of inside the gown anteriorly and both sides the whole length ,with two absorbant towels,		
		5. One reinforced surgical gowns(XL) with entire full reinforcement of inside the gown anteriorly and both sides the whole length with one absorbant towels		
		6. One cardio vascular top drape with fabric reinforcement(140-155cm)X(225-245CM) with positioning tape and adhesive edge width at least 10cm with extra strong adhesive power.		
		7. One cardio vascular bottom drape with fabric reinforcement (150-165cm) X(240-260CM) with positioning tape and adhesive edge width at least10cm with extra strong adhesive power		
		8. Two cardiovascular side drape with fabric reinforcement(75-90cm)X(95-110CM) with positioning tape and adhesive edge width at least10cm with extra strong adhesive power with.one of the drape having perfusion tubing pockets (plastic) with clipping tapes		
		9. Two OR cloth towels blue color(high absorption)		
		10. One diathermy bag (34x40cm)		
		11. Iodine_imprignated Adhesive Incise Drape (35CMX35CM)		

---

### شروط العطاء

---

### شروط خاصة

---

1-Samples of the quoted items must be submitted along with the offers. Samples must show the same serial no., item no., as per tender documents specifications FOR EVALUATION PURPOSE, PURCHASE and RECEIPT purposes. RMS with this respect reserves the right to reject offers not in compliance with this term.

2-All goods are to be freshly prepared on dispatch/ delivery & should have at least 2/3 of its shelf life upon receipt. Any item carrying less than 2/3 of its shelf upon receipt should be accompanied with a confirmation that you accept to replace any remaining unused quantities at your expenses after the expiry date. Said confirmation is subject to the approval of the director of the Royal Medical Services and will incur a fine, which will be decided later by DRMS according to the loss that this discrepancy with terms of the tender has caused.

3-Goods must be dispatched/ delivered under the same storage conditions that comply with their nature; storage conditions must be mentioned clearly on all packages and relative documents.

4-Storage conditions as well as description of goods, lot no., origin of goods, expiry date and manufacturing date must be clearly indicated on all packages (inner and outer) and relative documents. If the manufacturing date is not applicable it must be provided on a separate document.

5-The delivered goods must be stamped, labeled or printed "sold to DRMS" Tender No. P1-2025-5, with irremovable ink.

6-All packages, parcels and related documents to be stamped or labeled with the following:DRMS/TENDER

P1-2025-5  
GHQ, JORDAN ARMED FORCES  
AMMAN – JORDAN

7-Custom clearance of goods shall be the responsibility of the Jordan Armed Forces (JAF); however, suppliers shall bear all costs incurred by handling charges and any demurrage or extra expenses incurred by the port's corporation or QAIA (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order,

delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).

8-Pricing must include services of sale, shipment, transportation, delivery from port to main medical stores.

9-The supplier or his local agent must furnish RMS with a guarantee stamped and legalized by the notary public equals to 115% of the total value of goods valid for (one year) after acceptance of goods.

10-The quality of the items should be according to accepted international standards, i.e. (FDA, CE, or equivalent standard), accompanied with quality certificates certified by JFDA.

For locally manufactured goods, bidder should submit a valid (GMP) certificate issued and certified by JFDA.

11-The fabric used in gowns and drapes should abide the European or American standards (Bidder should provide full comprehensive certificate for that).

13-For offers submitted in Jordanian Dinar, payment will be either by wire transfer or by bank check to be paid after receiving the documents from main medical stores. Any other way of payment will be rejected by the purchasing committee.

14-Bidders must submit their reservation/queries regarding tender specifications and/or special terms within the first half of the tender closing period starting from the tender announcement date, any reservation/queries submitted after the end of this period shall be rejected.

15-After arrival to main medical stores, unloading of goods will be the responsibility of the local agent.