

PS41 / 2024 / 92

مواصفات المعينات السمعية

2024

Behind the Ear Digital hearing aids

Behind the Ear Digital computer based fitting hearing aids with software and programming kit

Item No	Type	Fitting range to be covered	Hearing loss to be covered	Special Requires Features	Number of program	Quantity
1	BTE, Digital Minimum 12 channels not bands computer based programming	250-8000Hz	Moderate hearing loss 45 -65 dB HL	<ul style="list-style-type: none">• Digital noise canceller/reduction.• Feedback Manager system• Wide dynamic range digital signal processing• Built in tinnitus masker• Directional Microphone• Dynamic noise cancellation/spatial speech enhancement.• Speech enhancement.• Automatic listening environment detection.	5 or more programs	850

2	BTE, Digital Minimum 12 channels not bands computer based programming	250-8000Hz	Severe hearing loss 70- 90 dB HL	<ul style="list-style-type: none"> • Digital noise canceller/reduction. • Feedback Manager system • Wide dynamic range digital signal processing • Built in tinnitus masker • Directional Microphone • Dynamic noise cancellation/spatial speech enhancement. • Speech enhancement. • Automatic listening environment detection. 	5 or more programs	600
3	BTE, Digital Minimum 12 channels not bands computer based programming	250-8000Hz	Severe to profound hearing loss 95 dB HL UP to 120 dB HL	<ul style="list-style-type: none"> • Digital noise canceller/reduction. • Feedback Manager system • Wide dynamic range digital signal processing • Built in tinnitus masker • Directional Microphone • Dynamic noise cancellation/spatial speech enhancement. • Speech enhancement. • Automatic listening environment detection. 	5 or more programs	600

BTE for slopping hearing loss

4	BTE, Digital Minimum 12 channels not bands computer based programming	250-8000Hz	Slopping hearing loss from 40- 100 dB HL	<ul style="list-style-type: none"> • Digital noise canceller/reduction. • Feedback Manager system • Wide dynamic range digital signal processing • Built in tinnitus masker • Directional Microphone • Dynamic noise cancellation/spatial speech enhancement. • Speech enhancement. • Automatic listening environment detection. 	5 or more programs	500
---	--	------------	--	--	-----------------------	-----

3-8

Molded Digital Hearing Aids computer based fitting with software and programming kit

Item No	Type	Fitting range to be covered	Hearing loss to be covered	Special Requires Features	Number of program	Quantity
5	Complete In The Canal (CIC), Digital Minimum 12 channels not bands	250-8000Hz	40-65 dB HL	<ul style="list-style-type: none"> • Digital noise canceller/reduction. • Feedback Manager system • Wide dynamic range digital signal processing • Built in tinnitus masker • Directional Microphone • Dynamic noise cancellation/spatial speech enhancement. • Speech enhancement. • Automatic listening environment detection. 	5 or more programs	500
6	Complete In the Canal (CIC), Digital Minimum 12 channels not bands	250-8000Hz	70-90 dBHL	<ul style="list-style-type: none"> • Digital noise canceller/reduction. • Feedback Manager system • Wide dynamic range digital signal processing • Built in tinnitus masker • Directional Microphone • Dynamic noise cancellation/spatial speech enhancement. • Speech enhancement. • Automatic listening environment detection. 	5 or more programs	500

4-8

شروط العطاء الخاصة

- يجب ابراز شهادة التصنيع الاصلية للمواد المتقدمة بها الشركة المناقصة للعرض محتومة ومصدقة من غرفة التجارة والصناعة في بلد التصنيع عند تقديم العروض وقبل بدء الدراسة الفنية وفي حال عدم ابراز الشهادة لا ينظر في العروض المقدمة. ولن يتم دراستها فنيا.
- ان تكون المعينات السمعية المقدمة حاصلة على شهادة الـ FDA او MDR و ICE .
- ان تكون سنة الاصدار لموديل المعينات السمعية اقلها 3 سنوات من تاريخ العطاء.
- يحق للخدمات الطبية الملكية رفض اي نوع من المعينات السمعية تم طرحها في السابق ولم تكن بالمستوى المطلوب او تم تقديم شكاوي ضدها او ضد الوكيل المناقص في الخدمات الطبية الملكية .
- ان يكون المناقص اردني الجنسية ولديه مزاولة مهنة في مجال السمعيات.
- تقديم شهادة ترخيص بيع معينات سمعية سارية المفعول ومزاولة مهنة للشخص الذي يقوم بالتركيب والمعايرة والصيانة من وزارة الصحة الاردنية .
- ان تكون مراجعة المريض لنفس المركز او الشركة المحال عليها ، ولا يجوز ارسال المريض الى مركز او شركة اخرى غير تابعة للمركز او الشركة المحال عليها.
- تسليم عينة من كل نوع من انواع المعينات السمعية المقدمة بالعرض لمديرية المشتريات العامة عند تقديم العروض كي يتم فحصها من قبل اللجنة الفنية ، وبعد الاحالة يتم تسليم المعينة السمعية المطابقة الموجودة في مديرية المشتريات لقسم السمعيات في مدينة الحسين الطبية.
- يجب تحديد نوع المعينة السمعية المخصصة لكل بند مبينا عليها الاسم، النوع والموديل و ان يكون تاريخ الصنع للسماعة بنفس العام المحال فيه العطاء.
- المواصفات المذكورة اعلاه هي الحد الأدنى من المواصفات وفي حالة توفر مواصفات إضافية في السماعة بالإضافة إلى المواصفات المذكورة تحسب ايجاباً لصالح الشركة عند فتح العطاء.
- تعطى اثنتا عشر بطارية مجانية مع كل سماعة نوع Zinc Air ، او ان تكون المعينة السمعية قابلة لإعادة الشحن.

- لا يتم اعتبار المعينة السمعية مطابقة للمواصفات الفنية المطلوبة الا بعد التأكد من اجراء الفحص الفني في قسم السمعيات في مدينة الحسين الطبية بغض النظر انها مطابقة للكتالوجات الاصلية .
- يتم فحص جميع السماعات عند دراسة المواصفات على حساب (2 CC Coupler) ويجب أن تغطي اعلى درجات ضعف السمع وعلى جميع الذبذبات المطلوبة وان تكون نسبة التشويش الداخلي (Harmonic Distortion) اقل ما يمكن .
- يجب ان تكون حجرة البطارية غير قابلة للكسر.
- عند الاستلام يجب أن تكون مطابقة ل (The current ANSI S3.22 for 2cc coupler Standard) وسيتم التأكد من ذلك من خلال فحصها في الخدمات الطبية وفي حالة عدم مطابقة أي سماعة يجب على الوكيل استبدالها بأخرى مطابقة خلال مدة أسبوعين من إعلام الشركة.
- تقديم كفاءة صيانة بنكية بقيمة 5% من اجمالي الاحالة لمدة عامين من تاريخ الاستلام النهائي للمعينات السمعية في الخدمات الطبية الملكية.
- تقديم كفاءة صيانة خطية مجانية من الشركة لكل مريض تحمل الرقم التسلسلي للمعينة السمعية لمدة عامين من تاريخ استلام المريض للمعينة السمعية.
- تقديم تعهد بالصيانة لمدة ثلاث سنوات بعد انتهاء كفاءة الصيانة البنكية المجانية و على حساب المريض.
- يتعهد الوكيل باجراء الصيانة اللازمة للمعينات السمعية المعطلة خلال فترة الكفاءة بمدة اقصاها عشرة ايام من تاريخ تسليمها لأحد فروع الشركة.
- يتعهد الوكيل بتثبيت اسعار قطع الغيار لمدة 5 سنوات من تاريخ انتهاء فترة الكفاءة المجانية بحيث يسمح ب 2% زيادة سنوية على الاسعار.
- يتم تسليم قالب أذن من نوع سيليكون او حسب رأي اخصائي السمعيات لكل مريض ولكل سماعة في موعد أقصاه اسبوعين من تاريخ تسليم المعينة السمعية لأحد فروع الشركة ,وان يلتزم الوكيل بتغيير القالب اذا ثبت انه غير ملائم وحسب الطلب الموقع من اخصاصي السمعيات.

Special Terms

1. *A. DRMS has right to increase the awarded quantities by a percentage not exceeding 35% after final order notification with the same prices, terms & conditions of the contract upon request & approval of awarded party .*

B. DRMS has right to decrease the awarded quantities by a percentage not exceeding 50% after final order notification with the same prices, terms & conditions of the contract upon request & approval of awarded party.
2. *Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).*
3. *For offers submitted in Jordanian dinars , payment will be either by wire transfer or by cheque after final acceptance of good , any other way of payment will be rejected.*
4. *Prices for all awarded items should be fixed for (3years)after receipt of final order.*
5. *The supplier must furnish DRMS with a guarantee letter stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded items valid for twelve months from the date of final acceptance of the equipment by DRMS which shall cover rusting, manufacturing defects, wear and tear and corrosion.*

6. Bidder must submit their reservations /queries regarding tender specifications and /or special terms within the first half of the tender closing period starting from the tender announcement date . Reservations/ queries submitted after the end of this period shall be rejected.

B-8