

FOLDABLE LENSES

TENDER NO P17-2023-B-8

| Sl. No. | Particulars | Qty | Unit | Remarks |
|---------|-------------|-----|------|---------|
|---------|-------------|-----|------|---------|

Foldable Intraocular Lenses
 Intraocular lenses 1-I.O.L Material: Acrylic with U.V
 and blue light filtering chromophore (Methylphenylazo
 hydroxyphenyl Ethylmethacrylamide):
 1.Three pieces with stableforce haptic
 2.Aspheric design 3A/constant :116.5-119.5
 4.Injectable including cartilage and injector in 3.0mm
 or less incision

| | | | |
|----|---------------------|--------|-----------------------|
| 1 | 20_14_00313_05 0 | 2 EACH | Diopter diameter 5.0 |
| 2 | 20_14_00313_05 5 | 2 EACH | Diopter diameter 5.5 |
| 3 | 20_14_00313_06 0 | 2 EACH | Diopter diameter 6.0 |
| 4 | 20_14_00313_06 5 | 2 EACH | Diopter diameter 6.5 |
| 5 | 20_14_00313_07 0 | 2 EACH | Diopter diameter 7 |
| 6 | 20_14_00313_07 5 | 2 EACH | Diopter diameter 7.5 |
| 7 | 20_14_00313_08 0 | 2 EACH | Diopter diameter 8.0 |
| 8 | 20_14_00313_08 5 | 2 EACH | Diopter diameter 8.5 |
| 9 | 20_14_00313_09 0 | 2 EACH | Diopter diameter 9.0 |
| 10 | 20_14_00313_09 5 | 2 EACH | Diopter diameter 9.5 |
| 11 | 20_14_00313_10 0 | 4 EACH | Diopter diameter 10.0 |
| 12 | 20_14_00313_10 5 | 2 EACH | Diopter diameter 10.5 |
| 13 | 20_14_00313_11 0 | 4 EACH | Diopter diameter 11.0 |
| 14 | 20_14_00313_11 5 | 2 EACH | Diopter diameter 11.5 |
| 15 | 20_14_00313_12 0 | 2 EACH | Diopter diameter 12.0 |
| 16 | 20_14_00313_12 5 | 4 EACH | Diopter diameter 12.5 |
| 17 | 20_14_00313_13 0 | 2 EACH | Diopter diameter 13.0 |
| 18 | 20_14_00313_13 5 | 4 EACH | Diopter diameter 13.5 |

FOLDABLE LENSES

TENDER NO P17-2023-B-8

| ITEM NO. QUANTITY UNIT DESCRIPTION | | | |
|------------------------------------|---------------------|----------|------------------------|
| 19 | 20_14_00313_14 0 | 2 EACH | Dioppter diameter 14.0 |
| 20 | 20_14_00313_14 5 | 4 EACH | Dioppter diameter 14.5 |
| 21 | 20_14_00313_15 0 | 4 EACH | Dioppter diameter 15.0 |
| 22 | 20_14_00313_15 5 | 8 EACH | Dioppter diameter 15.5 |
| 23 | 20_14_00313_16 0 | 8 EACH | Dioppter diameter 16.0 |
| 24 | 20_14_00313_16 5 | 8 EACH | Dioppter diameter 16.5 |
| 25 | 20_14_00313_17 0 | 8 EACH | Dioppter diameter 17.0 |
| 26 | 20_14_00313_17 5 | 12 EACH | Dioppter diameter 17.5 |
| 27 | 20_14_00313_18 0 | 76 EACH | Dioppter diameter 18.0 |
| 28 | 20_14_00313_18 5 | 76 EACH | Dioppter diameter 18.5 |
| 29 | 20_14_00313_19 0 | 84 EACH | Dioppter diameter 19.0 |
| 30 | 20_14_00313_19 5 | 84 EACH | Dioppter diameter 19.5 |
| 31 | 20_14_00313_20 0 | 108 EACH | Dioppter diameter 20.0 |
| 32 | 20_14_00313_20 5 | 112 EACH | Dioppter diameter 20.5 |
| 33 | 20_14_00313_21 0 | 108 EACH | Dioppter diameter 21.0 |
| 34 | 20_14_00313_21 5 | 112 EACH | Dioppter diameter 21.5 |
| 35 | 20_14_00313_22 0 | 108 EACH | Dioppter diameter 22.0 |
| 36 | 20_14_00313_22 5 | 112 EACH | Dioppter diameter 22.5 |
| 37 | 20_14_00313_23 0 | 108 EACH | Dioppter diameter 23.0 |
| 38 | 20_14_00313_23 5 | 112 EACH | Dioppter diameter 23.5 |
| 39 | 20_14_00313_24 0 | 108 EACH | Dioppter diameter 24.0 |
| 40 | 20_14_00313_24 5 | 10 EACH | Dioppter diameter 24.5 |

DIRECTORATE ROYAL MEDICAL SERVICES

FOLDABLE LENSES

TENDER NO P17-2023-B-8

| ITEM NO | DESCRIPTION | QUANTITY | UNIT | REMARKS |
|---------|---------------------|----------|------|------------------------|
| 41 | 20_14_00313_25 0 | 108 | EACH | Dioppter diameter 25.0 |
| 42 | 20_14_00313_25 5 | 66 | EACH | Dioppter diameter 25.5 |
| 43 | 20_14_00313_26 0 | 40 | EACH | Dioppter diameter 26.0 |
| 44 | 20_14_00313_26 5 | 24 | EACH | Dioppter diameter 26.5 |
| 45 | 20_14_00313_27 0 | 20 | EACH | Dioppter diameter 27.0 |
| 46 | 20_14_00313_27 5 | 12 | EACH | Dioppter diameter 27.5 |
| 47 | 20_14_00313_28 0 | 8 | EACH | Dioppter diameter 28.0 |
| 48 | 20_14_00313_28 5 | 8 | EACH | Dioppter diameter 28.5 |
| 49 | 20_14_00313_29 0 | 8 | EACH | Dioppter diameter 29.0 |
| 50 | 20_14_00313_29 5 | 4 | EACH | Dioppter diameter 29.5 |
| 51 | 20_14_00313_30 0 | 4 | EACH | Dioppter diameter 30.0 |

FOLDABLE LENSES P17-2023-B-8 الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

- 1 – Goods including injectors must be freshly prepared on dispatch/delivery and must retain at least 2/3 of its shelf life when dispatched /delivered. The DRMS with this respect reserves the right to reject any item not in compliance with this term and impose the correspondent fines as per GHQ /Military procurement laws .
- 2 – Dispatched/delivered goods must show description of goods, manufacturing date, expiry date, lot number, storage conditions and any other necessary information. Where the manufacturing date is not applicable, it should be provided on a separate document.
- 3 – Goods must be dispatch under the same storage conditions that comply with their nature & Storage Conditions Must be Clearly Indicated on Packages as Well as in the Relevant Documents.
- 4 – Quoted lenses should be previously approved by DRMS, and supportive documents should be provided with the offers.
- 5 – DRMS has the right to exchange lenses damaged during surgical operations if it is due to manufacturing default.
- 6 – DRMS has the right to exchange any power of the awarded lenses with any other power through the local agent, such exchange should be affected within a week after receipt of our written request. Any delay will impose penalties as per GHQ/RMS regulations.

FOLDABLE LENSES P17-2023-B-8 الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

- 7 - After arrival to Jordan, DRMS is not responsible for any handling charges for goods (like charges for issuing delivery note vouchers, unloading charges, local shipping agents charges). Furthermore, DRMS is not responsible for any demurrage charges and other charges incurred by the port corporation caused by delay in presenting all necessary shipping documents for clearing of goods.
- 8 - Shipment/ delivery: in Two Shipments/deliveries (4-5) months apart.
- 9 - All parcels and packages and related documents should be marked:
DRMS, tender no. P17-2023-B-8
GHQ/ Jordan Armed Forces, Amman, Jordan.
- 10 - The local agent should has all powers from 5 to 30 dioptr with increasing power 0.5
- 11 - Prices must include services of sale, shipment, transportation, delivery from port to site or Main Medical Stores.
- 12 - Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), how ever, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation or QAIA (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges) .The supplier is also responsible for providing af all relevant shipping documents ,together with the delivery orders.

FOLDABLE LENSES P17-2023-B-8 الشروط الخاصة للعطاء رقم

الشروط الخاصة التسلسل

- 13 - SAMPLES AND CATALOGUES OF THE QUOTED ITEM MUST BE SUBMITTED ALONG WITH THE OFFERS FOR PURCHASING AND RECEIPT PURPOSES SHOWING THE SAME SERIAL NUMBER/ CAT NUMBER AS PER OUR TENDER DOCUMENTS SPECIFICATIONS .RMS WITH THIS RESERVES THE RIGHT TO REJECT OFFERS NOT IN COMPLIANCE WITH THIS TERM.
- 14 - Reusable injectors (Qty:5each) should be submmited on F.O.C basis.
- 15 - In case of the required dioptr, A-constant and optic size asked to be exchanged is not provided, DRMS with this regard has the right to buy the suitable lenses from any other sources at your expenses and will incur the correspondent fine as per GHQ/ military procurement law.
- 16 - for offers submitted in jordanian dinars payment will be either by wire transfer or by cheque to be paid after receiving goods any other way of payment will not be accepted and the offers will be rejected by the pur chase committee.
- 17 - The supplier should state the estimated price for F.O.C goods