

JORDANIAN ARMY FORCES
Directorate Royal Medical
Services



القيادة العامة للقوات المسلحة الأردنية - الجيش العربي
مديرية الخدمات الطبية الملكية
شعبة المشتريات المركزيّة

الرقم : ش. ٢٤/٢٠٢٤/٦٠٠ / A
التاريخ : ربيع الثاني / ٢٤٤٦
: تشرين الثاني / ٢٠٢٤

السادة : المناقصون
الموضوع : المواصفات

تحية وبعد ،،،

اشارة الى العطاء رقم ش. ٢٤/٢٠٢٤/٦٠٠ / A
(نواقص احتياج وحدات الخدمات الطبية الملكية من عدد الاسنان)

- ١- مرفق طيه التعديل على مواصفات العطاء الاشارة أعلاه .
- ٢- لاجراءاتكم لطفاً علماً بأن لا تعديل على موعد اغلاق العطاء.

وتفضلوا بقبول فائق الاحترام

ع / العميد الطبيب
مدير عام الخدمات الطبية الملكية
عقيد سيدلاني طارق محمد الجبوري

A / ٢٠٢٤ | ٦٠٠ ٢

١. تعدل مواصفة المواد ارقام (٢٤, ٢٥, ٢٦) كما هو مبين بالجدول ادناه :

المواصفة بعد التعديل		المواصفة قبل التعديل	رقم المادة
50 pc	Cushing Tissue Forceps 17.5CM 1X2 without tooth	Cushing Tissue Forceps 17.5CM 1X2 Teeth	٢٤
50 pc	Cushing Tissue Forceps 17.5CM 1X2 with tooth		
Obwegeser HM Wire Twister 12-14CM		Obwegeser HM Wire Twister 12.5CM	٢٥
BAND PUSHER MERSHON ANGLED (Used for positioning and adapting metal bands on teeth.).		BAND PUSHER MERSHON. (Used for positioning and adapting metal bands on teeth.).	٢٦

٣. يعدل الشرط رقم (4/B) من الشروط الخاصة العائدة للعطاء اعلاه ليصبح

- For instruments of other origins, a copy of either a CE certificate (MDD)/TUV
OR a certificate of FDA approval OR MDR CE certificate

بدلا من

- For instruments of other origins, a copy of either a CE certificate (MDD)/TUV OR a
certificate of FDA approval

Item No	Stock No	Name of Instrument	unit	Required Amount
General Dentistry				
1.	60_11_103	Alginate mixing Spatula, PLASTIC	Each	80
2.	60_11_104	Alginate Mixing Bowl.	Each	30
3.	60_11_202/L	Dental Sterilization Box Large Size.	Each	50
4.	60_11_202/M	Dental Sterilization Box Medium Size.	Each	50
5.	60_11_306	Waste Receiver	Each	50
6.	60_11_401	Dental Anesthesia Syringe.	Each	200
Conservative and Endodontics				
7.	60_12_204/N	Amalgam Burnishes Narrow Blunt Ended.	Each	50
8.	60_12_408	Rubber Dam Complete Kit	kit	20
9.	New	MTA Carrier	Each	10
10.	New	MTA micro apical application system	kit	10
11.	New	Micro condensers , kit contain assorted shapes & sizes for root-end filling.	kit	10
Periodontics				
12.	60_13_201/S	Curette Surgical Curette Single End.	Each	20
13.	60_13_304/L	TWEEZER CRANE-KAPLAN (POCKET MARKER) UPPER LEFT (LOWER RIGHT)	Each	50
14.	60_13_304/R	TWEEZER CRANE-KAPLAN (POCKET MARKER) UPPER RIGHT(LOWER LEFT)	Each	50
15.	60_13_401	SHARPENING STONE	Each	30
16.	60_13_807	Dental Bone Filter	Each	10
17.	60_13_808	Dental Bone Crusher	Each	10
Prosthodontics				
18.	60_14_603/A	Dental Bridge Removal Automatic	Each	25
19.	60_14_605	Dental Laboratory Crown holder	Each	10
Maxillofacial Surgery				
20.	60_15_303	Surgical Wire Cutter	Each	30
21.	60_15_308	Dental Blade Handle	Each	150
22.	60_15_314	Dental Blade Remover	Each	100
23.	60_15_317	Dental Plate Cutter	Each	50
24.	60_15_318/T	Cushing Tissue Forceps 17.5CM 1X2 Teeth	Each	100
25.	60_15_402	Obwegeser HM Wire Twister 12.5CM	Each	50
Orthodontics				
26.	60_17_001	BAND PUSHER MERSHON. (Used for positioning and adapting metal bands on teeth.).	Each	70
27.	60_17_204	BRACKET REMOVER PLIER.	Each	50
28.	60_17_507	Distal End Cutter (Long) (Used to cut orthodontic wires as close as possible to the buccal tube, holding securely the cut wire portion, max 0, 55 x 0, 64 mm.)	Each	100
29.	60_17_726/2	Orthodontic Torqueing Plier 0.022"	Each	30
30.	60_17_802	Tension and compression gauge for measuring orthodontic forces	Each	30



Special terms for medical instruments

- Offers not complying with any of the special terms or the technical specifications shall be considered non-conforming with tender requirements.
- Any vendor providing FORGED documents shall be disqualified from the current tender and banned from participating in any future RMS tenders
- Offers submitted by vendors with previously complaints or unsolved issues will be excluded from the current tender

1. The vendor is responsible to ensure through official documents that classified medical instrument are manufactured in conformity with applicable quality system standards (ISO) and must provide the latest copy of the certificates of the following standards formally endorsed by the chamber of commerce of the manufacturing origin with the technical offer:

A- ISO 13485:2016

Medical devices — Quality management systems — Requirements for regulatory purposes

ISO 13485:2016 specifies requirements for a quality management system where an organization needs to demonstrate its ability to provide medical devices and related services that consistently meet customer and applicable regulatory requirements.



Special terms for medical instruments

B- ISO 9001:2015

Quality management systems — Requirements

specifies requirements for a quality management system aims to enhance customer satisfaction through the effective application of the system, including processes for improvement of the system and the assurance of conformity to customer and applicable statutory and regulatory requirements.

C- ISO 7153-1:2016

Surgical instruments — Materials — Part 1: Metals

specifies metals commonly used to manufacture various types of standard surgical instruments, including but not limited to those used in general surgery, orthopaedics and dentistry.

D- ISO 7151:1988

Surgical instruments — Non-cutting, articulated instruments — General requirements and test methods

Specification of basic requirements for as well physical characteristics as workmanship, of the steel grades used and heat treatment of component parts, excluding rivets, screws and parts manufactured of material grade M.

E- ISO 7740:1985

Instruments for surgery — Scalpels with detachable blades — Fitting dimensions

Lays down the dimensions of two sizes of fitting features for detachable scalpel blades and the handles with which they are used. It secures a good fitting and interchangeability of detachable blades for scalpels manufactured in different countries or by different manufacturers



Special terms for medical instruments

F- ISO 7741:1986

Instruments for surgery — Scissors and shears — General requirements and test methods

This standard deals with materials, heat treatment and hardness of component parts, corrosion resistance, workmanship and cutting ability of scissors and shears used in the surgery and defines the test methods.

G- ISO 13402:1995

Surgical and dental hand instruments — Determination of resistance against autoclaving, corrosion and thermal exposure

Describes test methods to determine the resistance of stainless steel surgical and dental hand instruments against autoclaving, corrosion and thermal exposure.

2. Offered items should be from reputable well known manufacturers and excellent experience in the field previously evaluated and accepted in Main Hospitals at Royal Medical Service . Bidder must submit samples with the technical offer from each brand that was submitted in the offer to be assessed and /or evaluated by RMS
 - A. Evaluation duration and quantities will be determined by the purchasing committee.
 - B. Offers which do not include such samples will be considered non-conforming.
 - C. Offers which fail the evaluation/assessment process will be excluded from the tender.
 - D. Samples will be returned to the Bidder at the end of the evaluation/assessment process except for samples from Bidders winning the final award where they will be the property of RMS and will be kept as a reference for the receiving committee.



Special terms for medical instruments

- E. The purchasing committee has the right to require a sample for any instrument previously installed in RMS Hospitals.
3. A. Bidder is not allowed to submit more than one offer for the same item whether that's solely or in coalition or partnership with other bidders.
B. Bidder is allowed to include within their offer optional alternatives for the same offered item from the same manufacturer.
4. Required certificates:
 - A. For instruments of USA origin, a copy of a certificate of FDA approval must be submitted with the technical offer.
 - B. For instruments of other origins, a copy of either a CE certificate (MDD)/TUV OR a certificate of FDA approval must be submitted with the technical offer.
 - C. With each offer, bidders must provide a formally endorsed document issued by the manufacturer stating that the bidder is the authorized agent/ distributor for the offered item.
 - D. In all of the above cases (except C) certificates must be formally endorsed by JFDA .
 - E. Any vendor not submitting all required certificates will be eliminated.
5. All Offered items should be supplied with free sale certificate said confirmation should be stating clearly that the item is freely sold in the country of origin .
6. All items should be engraved or etched with manufacturing origin ,company logo and code number .
7. Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first third of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period will be rejected



Special terms for medical instruments

- date. Reservations/queries submitted after the end of this period will be rejected
9. The supplier must furnish DRMS with a guarantee letter stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded equipment valid for (24) months from the date of final acceptance of the instruments by DRMS which shall cover rusting, manufacturing defects, wear and tear and corrosion.
 10. Offers must include clear original catalogues for all offered items. Images and part numbers of offered items must be provided, highlighted and outlined clearly.
 10. Offers must include fully detailed information as a soft copy (either Microsoft office or Microsoft excel format) in addition to a hard copy, mentioning the exact model/catalogue number and country of origin of the offered item(s), full description/specifications. And any accessories or options included in the offer.
 11. Official website of the manufacturer must be provided. The website must clearly demonstrate the history profile and manufacturing features of the company. In case of inconsistency between the information on the website with the provided information of the local agent regarding any offered product, the purchasing committee has the right to eliminate that product from the awarding process.
 12. Any complementary parts, accessories and consumable items necessary for the proper operation of surgical Instruments must be included in the offer, priced separately and to be be approved by the manufacturer.
 13. A. complementary parts availability must be guaranteed and can be purchased without the need replace the entire system of surgical instruments
B. Bidder will bear any responsibility resulting from the awarding of insufficient items or sub-items.
 14. Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores.
 15. A. custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all



Special terms for medical instruments

costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.).

B. The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).

16. Delivery period should be mentioned clearly; the purchasing committee has the right to reject any offer with delivery time exceeding 4 months.
17. Goods are to be brand new; manufacturing date of the delivered products should not exceed 18 months from the date of the final award.
18. A. DRMS has the right to increase the awarded quantities by a percentage not exceeding 35% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.
B. DRMS has the right to decrease the awarded quantities by a percentage not exceeding 50% after the final order notification with the same prices, terms and conditions of the original contract upon DRMS request and approval of the awarded party.

19 . For offers submitted in Jordanian dinars, payment will be either by wire transfer or by cheque after final acceptance of goods, Any other way of payment will be rejected .