



### محضر اجتماع

اجتمعت اللجنة المشكلة بموجب كتاب عطوفة مدير عام الخدمات الطبية الملكية رقم ش ٣٥٧٧٥/١١/٢٠٢٤/٧٠٠ تاريخ ٢٠٢٤/١١/٧ لدراسة احتياج الخدمات الطبية الملكية من مستلزمات جراحة الوجه و الفكين لمدة عام حيث توصي اللجنة بالتعديلات التالية على المواصفات المطروحة حاليا في شعبة المشتريات المركزية :

#### - المادة رقم (Item no.2) يصبح

Microplate		
#	Description	QTY
2.1	Superior orbital rim	20
2.2	Inferior orbital rim	50
2.3	Straight	50
2.4	Double Y	30

بدلا من

Microplate		
#	Description	QTY
2	Microplate	150

#### - المادة رقم (Item no.4) يصبح الوصف

2.0 mmMP mini plates

بدلا من

2.5 mmMP mini plates

#### - تلغى المواد التالية:

- Item no.5.3
- Item no 6.6
- Item no 6.7
- Item no 6.8
- Item no 6.9
- Item no 6.10

#### - المادة رقم (Item no.9) يصبح الوصف

2.0 mm Bone screws (self drilling)

بدلا من

2.0 mm Bone screws



تضاف البنود التالية للشروط رقم (24):

- Twist drill bits for microplates (minimum 5 pieces).
- (5) screw drivers with their blades for microplates screw.

### **Item no.1**

2.0 mm Bone Screws

2.0 mm MP self-tapping		
#	Description	QTY
1.1	2.0*6 mm	1500
1.2	2.0*8 mm	1500
1.3	2.0*10 mm	1000
1.4	2.0*12 mm	500
1.5	1.5*6 mm	50
1.6	1.5*8 mm	50

### **Item no.2**

Microplate

Microplate		
#	Description	QTY
2	Microplate	150

### **Item no.3**

2.3 mm Bone Screws (Emergency)

2.3 mm MP self-tapping (Emergency)		
#	Description	QTY
3.1	2.3*6 mm	100
3.2	2.3*8 mm	100
3.3	2.3*10 mm	200
3.4	2.3*12 mm	100

#### **Item no.4**

2.5 mm MP mini plates

#	Description	QTY
4.1	Straight plates 16 holes regular	150
4.2	L-plate 8mm bar , 5 holes regular -R	100
4.3	L-plate 4mm bar , 4 holes regular -R	50
4.4	L-plate 8mm bar , 5 holes regular -L	100
4.5	L-plate 4mm bar , 4 holes regular -L	50
4.6	CHIN plates 4mm bar	5
4.7	CHIN plates 6mm bar	5
4.8	CHIN plates 8mm bar	5
4.9	Y-plate	50
4.10	Double Y-plate	50

#### **Item no.5**

MMF screws

#	Description	QTY
5.1	2.0mm MMF self drilling screws 8 mm	100
5.2	2.0mm MMF self drilling screws 12 mm	100
5.3	Hybrid MMF	10

#### **Item no.6**

Orbital floor plates 0.3 mm

#	Description	QTY
6.1	Isolated small	10
6.2	Basic medium	30
6.3	Complex large	5
6.4	Titanium orbital mesh 0.6 mm	5
6.5	Titanium orbital mesh 0.3 mm	7
6.6	Medpore orbital floor	3
6.7	Orbital rim plates 0.6mm, 8 holes	10
6.8	Resorbable orbital implant 30*30 mm, 0.2mm	1
6.9	Resorbable orbital implant 25*25 mm, 0.1 mm	1
6.10	Resorbable orbital mesh, 0.6 mm	1

### **Item no.7**

Mandibular reconstruction plate

#	Description	QTY
7.1	Mandibular reconstruction plate, hemi, right side with template	15
7.2	Mandibular reconstruction plate, hemi, left side with template	15
7.3	Low profile reconstruction plate	15

### **Item no.8**

Temporary condylar prosthesis

#	Description	QTY
8.1	Temporary condylar prosthesis, left side for universal mandible system	2
8.2	Temporary condylar prosthesis, right side for universal mandible system	2

### **Item no.9**

2.0 mm Bone Screws

#	Description	QTY
9.1	2.0*6 mm	250
9.2	2.0*8 mm	250

Jordanian Armed Forces – Arab Army



Directorate of Royal Medical Services  
Directorate of Pharmacy and Medical Supply  
Special Terms for Implant Tools (Plates, Screws, & Prosthesis)

1. *Offers not complying with any of the special terms or the technical specifications will be considered as non-conforming.*
2. *Any vendor providing FORGED documents will be disqualified from the current tender and any future RMS tenders or purchase orders.*
3. *Required certificates:*
  - a. *For Implant tools of USA origin, a copy of a certificate of FDA approval must be submitted with the technical offer.*
  - b. *For implant tools of other origins, a copy of either a CE certificate , MDR certificate, TÜV OR a certificate of FDA approval must be submitted with the technical offer.*
  - c. *With each offer, bidders must provide a formally endorsed document issued by the manufacturer stating that the bidder is the sole certified agent / distributor for the offered item.*
  - d. *In all of the above cases (except C) certificates must be formally endorsed by JFDA.*
  - e. *Regarding Joint prosthesis there must be a good rank by ODEP (Orthopedic Data Evaluation Panel).*
4. *Implants must be previously evaluated and approved by DRMS or purchased by tenders or contracts.*
5. *Bidders are obligated to provide RMS with well recognised international hospitals (centers) to which their products are being distributed and used without any recall or complaints .*
6. *Goods are to be brand new and of the latest version from the manufacturer company*
7. *Goods must be freshly prepared on dispatch and must retain at least 2/3 of its shelf life when dispatch where is applicable, DRMS has the right to reject any item not in compliance with this term and impose the correspondent fines as per DRMS procurement laws*
8. *Bidders must submit samples with the technical offer from each brand that was submitted in the offer to be assessed and /or evaluated by RMS.*
  - *Offers which do not include such samples will be considered non-conforming.*
  - *Offers which fail the assessment process will be excluded from the tender.*
  - *Samples will be returned to the bidder at the end of the assessment process except for samples from bidders winning the final award where they will be the property of RMS and will be kept as a reference for the receiving committee.*
9. *Prices should be fixed for (3 years) after receipt of final order.*
10. *All items should be engraved or etched with manufacturing origin, company logo and code number (where applicable).*
11. *Bidders are obliged to provide RMS with brand new surgical instruments needed for operation on a case by case basis until all purchased quantities have been used in any RMS hospital.*
12. *Bidders are obliged to exchange the awarded item with any other size according to RMS needs until all purchased quantities have been used. Such exchange should be within (one week) after receipt of a written request from RMS*
13. *Goods should be dispatched under the same storage conditions that comply with their nature, storage conditions must be mentioned clearly on all documents.*

Jordanian Armed Forces – Arab Army



Directorate of Royal Medical Services  
Directorate of Pharmacy and Medical Supply  
Special Terms for Implant Tools (Plates, Screws, & Prosthesis)

14. *A. DRMS has right to increase the awarded quantities by a percentage not exceeding 35% after final order notification with the same prices, terms & conditions of the contract upon request & approval of awarded party.*  
*B. DRMS has right to decrease the awarded quantities by a percentage not exceeding 50% after final order notification with the same prices, terms & conditions of the contract upon request & approval of awarded party.*
15. *Local agents are obliged to send an operation technician to participate in the surgical operations at RMS hospital until all purchased quantities have been used.*
16. *In the case where Any defect or claim of any item is found during surgical operations, the item should be replaced by a new one within ( one week) after the receipt of a written request from RMS.*
17. *Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first half of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period will be rejected.*
18. *The supplier must furnish DRMS with a guarantee letter stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded items valid for twelve months from the date of final acceptance of the equipment by DRMS which shall cover rusting, manufacturing defects, wear and tear and corrosion.*
19. *Offers must include clear original catalogues for all offered items. Images and part numbers of offered items must be provided, highlighted and outlined clearly.*
20. *Offers must include fully detailed information as a soft copy (either Microsoft office or Microsoft excel format) in addition to a hard copy, mentioning the exact model/catalogue number and country of origin of the offered item(s), full description/specifications.*
21. *Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores.*
22. *Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).*
23. *For offers submitted in Jordanian dinars , payment will be either by wire transfer or by cheque after final acceptance of good , any other way of payment will be rejected.*

Jordanian Armed Forces – Arab Army



Directorate of Royal Medical Services  
Directorate of Pharmacy and Medical Supply  
Special Terms for Implant Tools (Plates, Screws, & Prosthesis)

24. Offers must include at least the following free of charge items:
- Ten (10) screw drivers with their blades.
  - Twist drill bits 1.5x50 mm, 5 mm stop (minimum 30 pieces)
  - Twist drill bits 1.5x50 mm, 7 mm stop (minimum 30 pieces)
  - Twist drill bits 1.5x50 mm, 20 mm stop (minimum 30 pieces)
  - Twist drill bits 1.5x50 mm, 20 mm stop (minimum 30 pieces)
  - Twist drill bits 1.9x85 mm, 35 mm stop (minimum 10 pieces) for reconstruction plate.
  - Twist drill bits 1.9x115 mm, 35 mm stop (minimum 10 pieces) for reconstruction plate.
25. Any other free of charge related surgical instruments (like plate bender, plate holder, plate cutters.....etc) will be accepted.
26. For item number (7) each unit must include (6)locking screws divided into (3) of diameter 2.3x10 mm and (3) of diameter 2.3x12 mm.
27. For item number (8) each unit must include (2) fixation screws.
28. For items number (7) & (8) bidders are obliged to exchange sub items with each other's according to DRMS needs at any time after completion purchasing process.
29. For items number (1.5)&(1.6) compatible (10) screw driver blades must be provided.
30. Offers must include (6) free mini kit screws.