



محضر اجتماع

الموضوع : العتاء رقم ش ٣٨/٢٠٢٢/٥٠٠ لشراء اجهزة طبية مساندة تصرف للمرضى .

اجتمعت لجنة مواصفات العطاء اعلاه و المشكلة بموجب كتب عطوفة مدير عام الخدمات الطبية الملكية رقم ش ١٨٦٨٤/٣٨/٢٠٢٢/٥٠٠ تاريخ ٢٠٢٢/٨/٤ و رقم ش ٢٤٠٨١/٣٨/٢٠٢٢/٥٠٠ تاريخ ٢٠٢٢/٩/٢٢ ذلك لدراسة الاستفسارات التالية :

• استفسار السادة شركة ايلا العلمية الرباعية رقم (أ ع ر / ٢٠/٢٣) تاريخ ٢٠٢٣/٣/١

• بخصوص المادة رقم (Item 13:Commode chair adult size) المواصفة رقم (٣) تعديل لتصبح :

Weight capacity ≥ 100 kg


بدلا من:

Weight capacity ≥ 135 kg

• لا تعديل على باقي المواصفات و الشروط الخاصة .



JORDAN ARMED FORCES



The DIRECTION OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 1	Wheel Chair, Adult	Qty. 125 ()
--------	--------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Adult size wheel-chair (seat width 42 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 2	Wheel Chair, Extra Wide Adult	Qty. (50)
--------	-------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Extra wide Adult size wheel-chair (seat width 56-60 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 3	Wheel Chair, Children	Qty. (20)
--------	-----------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Children size wheel-chair (seat width 36-38 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 4	Wheel Chair, Wide Adult	Qty. ¹⁷⁵ (200)
--------	-------------------------	--------------------------------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	wide Adult size wheel-chair (seat width 44-46 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 5	Wheel Chair, Narrow Adult	Qty. (30)
--------	---------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be rigid type		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	Armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	<i>Brochure Page No.</i>
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Narrow Adult size wheel-chair (seat width 38-40 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 6	WHEELCHAIR, WIDE ADULT (FULLY RECLINING BACK)	Qty. (10)
--------	-------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be tubeless or air free.		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair wide Adult size (seat width 45-46 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 7	WHEELCHAIR, ADULT(FULLY RECLINING BACK)	Qty. (10)
--------	-------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be tubeless or air free.		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Adult size (seat width 42 cm)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 8	WHEELCHAIR,NARRW ADULT(FULLY RECLINING BACK)	Qty. (10)
--------	------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be tubeless or air free.		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Adult size (seat width 38-40 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 9	WHEELCHAIR, CHILDREN (FULLY RECLINING BACK)	Qty. (30)
--------	-----------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
A	Wheels:		
1	The user propelled wheels should be 24 inches in diameter mounted into the rear of both side frames.		
2	Swivel casters (five to eight inches in diameter) should be mounted to the front of both side frames.		
3	The user propelled wheels should have a push rim attached.		
4	Each of the user-propelled wheels should have toggle or lever wheel locks		
5	Wheels shall be tubeless or air free.		
B	Seat and Backrest:		
1	The wheelchair should have collapsible seat and backrest.		
2	The wheelchair should have some type of armrest, and these armrests shall be adjustable in height and pivoting or removable for ease of transfers.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
3	The wheelchair should have leg rests or footrests, and they shall have adjustable lengths and either detachable or swing-away features for ease of transfer. Footrests should either flip-up or fold and should have impact guards.		
4	The wheelchair should have anti-tip devices and should not tip on inclines with slopes of up to 15°.		
5	The wheelchair should be easy to manoeuvre in confined spaces.		
6	The wheelchair upholstery should pass applicable standards for flame resistance.		
7	Fully reclining back(90-180) degree with head support wheel-chair Adult size (seat width 36-38 cm)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 10	WHEELCHAIR FOR CHILDREN WITH CEREBRAL PALSY(36cm)	Qty. (10)
---------	----------------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction aluminium frame.		
2	Fully reclining back (90-180) degree with head support wheel-chair (seat width 36 cm)		
3	Adjustable seat angle.		
4	Detachable desk arms rest		
5	Pneumatic tired with wheels for outdoor use (pump & English key are essential)		
6	Frame made for durable, anti-corrosion, easy to clean		
7	Durable front caster solid or pneumatic to be 360 pivot type, with locking mechanism		
8	Adjustable height head rest (for head stability)		
9	Adjustable height for plates		
10	Padded abductor for thigh separation		
11	Padding material should be durable, provides optimum comfort. easy to clean and disinfect.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 11	WHEELCHAIR FOR CHILDREN WITH CEREBRAL PALSY(30cm)	Qty. (25)
---------	----------------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction aluminium frame.		
2	Fully reclining back (90-180) degree with head support wheel-chair (seat width 30 cm)		
3	Adjustable seat angle.		
4	Detachable desk arms rest		
5	Pneumatic tired with wheels for outdoor use (pump & English key are essential)		
6	Frame made for durable, anti-corrosion, easy to clean		
7	Durable front caster solid or pneumatic to be 360 pivot type, with locking mechanism		
8	Adjustable height head rest (for head stability)		
9	Adjustable height for plates		
10	Padded abductor for thigh separation		
11	Padding material should be durable, provides optimum comfort, easy to clean and disinfect.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 12	Commode chair wide adult size (seat 46.5 cm)	Qty. (100)
---------	----------------------------------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.


Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size wheelchair self-propelled shower & commode chair (seat width 46.5cm).		
2	Flip up arm rest.		
3	Weight capacity \geq 135 kg		
4	Pneumatic tired nylon wheels with hand rims with(pump & English)		
5	Solid front castors wheels with lock		
6	detachable/swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 13	Commode chair adult size(seat 42cm)	Qty. (60)
---------	-------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size wheelchair self-propelled shower & commode chair (seat width 42cm).		
2	Flip up arm rest.		
3	Weight capacity \geq 135 kg		
4	Pneumatic tired nylon wheels with hand rims with(pump & English)		
5	Solid front castors wheels with lock		
6	Detachable/swinging foot rest.		
7	A gap fronted seat to allow its use as commode/ over toilet pedestal.		
8	Foldable chair frame for easy storage.		
9	Breaks for both sides.		
10	Galvanized metal as protection of rust		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 14	Walker, wide adult	Qty. (50)
---------	--------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Wide adult size walker. Approximately 95 cm height		
2	Adjustable height. 30cm – 40cm		
3	Aluminium bar.		
4	Foam hand grip.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 15	Walker, adult	Qty. (50)
---------	---------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Normal adult size walker Approximately 95 cm height		
2	Adjustable height. 30cm – 40cm		
3	Aluminium bar.		
4	Foam hand grip.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 16	Walker, children(2-7)	Qty. (20)
---------	-----------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Normal Children size walker. Approximately 45cm – 55cm		
2	Adjustable height.		
3	Aluminium bar.		
4	Foam hand grip.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 17	K-Walker, large	Qty. (20)
---------	-----------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 70cm – 80cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 18	K-Walker, small	Qty. ⁶⁰ (28)
---------	-----------------	------------------------------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 2 to 7.		
3	Adjustable height. 55cm – 65cm		
4	Aluminium bars.		
5	Nylon hand grip or plastic.		
6	Four small castors.		
7	Non-swivel front wheels.		
8	Two back castors with breaks.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 19	Two fronts Walker style children size(2-7)	Qty. (30)
---------	--------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 2 to 7.		
3	Adjustable height. 45cm – 55cm		
4	Aluminium bars.		
5	Padded hand grip.		
6	Two front wheels		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 20	Two fronts Walker style children size(7-12)	Qty. (40)
---------	---------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Heavy duty construction.		
2	Children size walker age from 7 to 12.		
3	Adjustable height. 80cm – 90cm		
4	Aluminium bars.		
5	padded hand grip.		
6	Two front wheels		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 21	Standing Frame style (supine/prone) medium size	Qty. (30)
---------	-------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size standing frame.		
2	For children 7-15 years.		
3	Motorized height adjustment of top .		
4	Unit can power lift up to around 140kgs patient from sitting to standing position.		
5	Hand control operation.		
6	top to be covered with a suitable		
7	Dimensions of top approximately (100X44) cm with raised rim.		
8	Cut-out size is approximately (48 W x 25 D) cm.		
9	Padded knee		
10	Chest and back supports holds patient in standing position.		
11	Heavy gauge steel powder coated frame.		
12	Retractable locking swivel casters.		
13	Frame base is approximately (94 W x 106 D) cm.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 22	Wheelchair pure gel medical cushion (43*41*3 cm)	Qty. (30)
---------	----------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion		
2	Height (3 cm)		
2	Cushion for adult size wheel-chair (43cm width * 41 length)		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 23	Wheelchair pure gel medical cushion (40*40*3 cm)	Qty. (40)
---------	--------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion.		
2	Height (3 cm.)		
2	Cushion for narrow adult size wheel-chair (40cm width * 40 cm length).		



JORDAN ARMED FORCES



The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 24	Wheelchair pure gel medical cushion (36*28*3 cm)	Qty. (25)
---------	--------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion		
2	Height (3 cm.)		
3	Cushion for children size wheel-chair (36cm width * 28cm length).		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 25	Wheelchair pure gel medical cushion (56*60*3 cm)	Qty. (25)
---------	--------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Pure gel (solid) medical cushion		
2	Height (3 cm.)		
3	Cushion for extra wide adult size wheel-chair (56cm width-60cm length).		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 26	Wheelchair foam medical(foam + gel) cushion(43*41*6 cm)	Qty. (30)
---------	----------------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	foam medical cushion(foam + gel)		
2	Height (1cm) gel , (5cm) foam.		
2	Cushion for adult size wheel-chair (43cm depth*41cm length)		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 27	Wheelchair foam medical(foam + gel) cushion(40*40*6 cm)	Qty. (30)
---------	----------------------------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	foam medical cushion(foam + gel)		
2	Height (1cm) gel , (5cm) foam.		
2	Cushion for adult size wheel-chair (40cm depth*40cm length)		



JORDAN ARMED FORCES



The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY

Item 28	Quadruped elbow crutches (Tall adult)	Qty. (30)
---------	-----------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Tall Adult size stick.130cm – 150cm height		
2	Strong, lightweight aluminium crutch.		
3	Tetrapod design has four high –grip feet give more stability.		
4	Adjustable handle height.		
5	Adjustable height.		
6	Padding hand grip		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 29	Quadruped elbow crutches (Adult)	Qty. (50)
---------	------------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size stick.110 – 130 cm height		
2	Strong, lightweight aluminium crutch.		
3	Tetrapod design has four high –grip feet give more stability.		
4	Adjustable handle height.		
5	Adjustable height.		
6	Padding hand grip		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 30	Patient stick , Quadruped cane	Qty. (55)
---------	--------------------------------	-------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Adult size stick.		
2	Aluminium shaft with chrome plated steel base.		
3	Adjustable quadruped stick.		
4	Moulded handle and plastic hand grip.		
5	Adjustable height approx. (70cm_95cm).		
6	Easily reverses for right or left hand.		



JORDAN ARMED FORCES

The DIRECTIONATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



Item 31	Standing Frame style (supine/prone) small size	Qty. (280)
---------	------------------------------------------------	--------------

IMPORTANT NOTE:

Vendors are required to fill out the product details, compliance information, and brochure reference page number in the columns below

TECHNICAL SPECIFICATIONS:

The unit must meet or exceed the requirements listed in the table below.

Product Details	
Name of Manufacturer	
Model/ catalogue number	
Country of Origin for the offered model	
Country where the manufacturer is based	
Delivery time	
Full warranty period	
FDA clearance OR CE Mark	

	Minimum Requirements	Compliance (Y/N), Notes	Brochure Page No.
1	Children size standing frame.		
2	For children 2-7 years.		
3	Motorized height adjustment of top .		
4	Unit can power lift up to around 140kgs patient from sitting to standing position.		
5	Hand control operation.		
6	top to be covered with a suitable		
7	Dimensions of top approximately (85X38) cm with raised rim.		
8	Cut-out size is approximately (48 W x 25 D) cm.		
9	Padded knee		
10	Chest and back supports holds patient in standing position.		
11	Heavy gauge steel powder coated frame.		
12	Retractable locking swivel casters.		
13	Frame base is approximately(94 W x 106 D) cm.		



JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



SPECIAL TERMS

- Offers not complying with any of the special terms or the technical specifications shall be considered non-conforming with tender requirements.
- Any vendor providing FORGED documents shall be disqualified from the current tender and banned from participating in any future RMS tenders.
 1. Bidders must submit their reservations/queries regarding tender specifications and/or special terms within the first third of the tender closing period starting from the tender announcement date. Reservations/queries submitted after the end of this period shall be rejected.
 2. Technical offers must include clear original technical brochures/catalogues for all offered items.
 3. Offers must include fully detailed technical offers and compliance sheets as a soft copy (either Microsoft office or Microsoft excel format) in addition to a hard copy, mentioning the exact model/catalogue number and country of origin of the offered item(s), full technical description/specifications and any accessories or options included in the offer.
 4. Compliance sheets must be as per the tabular format of the technical specifications in the tender documents, listing the required specifications on one column and a Yes or NO response to each point in the adjacent column, with reference to page and line numbers in the relevant technical brochure. Offers not complying with this term shall be rejected.
 5. Pricing must include services of sale, shipment, transportation, delivery from port to site or to Main Medical Stores and bringing the equipment into service.
 6. Custom clearance of goods shall be the responsibility of the Jordanian Armed Forces (JAF), however, suppliers shall bear all costs incurred by handling charges and any demurrage charges or extra expenses incurred by the port's corporation (including expenses caused by delay in presenting the necessary shipment documents for either clearing or transporting the goods to the required location mentioned in the final order, delivery note issuing charges, unloading charges, local shipping charges etc.). The supplier is also responsible for providing of all relevant shipping documents, together with the delivery order(s).

JORDAN ARMED FORCES

The DIRECTORATE OF ROYAL MEDICAL SERVICES
THE INSTITUTE OF BIOMEDICAL TECHNOLOGY



7. *The supplier must furnish DRMS with a guarantee stamped and legalized by the Notary Public equals to (115%) of the total value of the awarded equipment valid for twelvemonths from the date of final acceptance of the equipment by DRMS.*

8. For offers submitted in Jordanian dinars , payment will be either by wire transfer or by cheque after final acceptance of goods .Any other way of payment will be rejected .